



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE RD. SUITE 210
LISLE, IL 60532-4352

September 21, 2015

EN 51341
NMED No. 150481

Arthur J. Frazier, M.D.
Radiation Safety Officer
Mount Clemens Regional Medical Center
d/b/a McLaren Macomb
1000 Harrington Blvd.
Mount Clemens, MI 48043

SUBJECT: ACKNOWLEDGMENT OF LETTER PROVIDING NOTIFICATION OF A
REPORTABLE EVENT DATED SEPTEMBER 9, 2015 – MOUNT CLEMENS
REGIONAL MEDICAL CENTER D/B/A MCLAREN MACOMB

Dear Dr. Frazier:

This acknowledges receipt of your letter dated September 9, 2015, reporting an incident that occurred on August 18, 2015, which involved an iridium-192 sealed source that got stuck in a source position check ruler during acceptance testing. The NRC has reviewed your letter and its attachments and has no further questions at this time. The incident will be examined during a future inspection.

In accordance with Title 10 of the *Code of Federal Regulations* (CFR) 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC's Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC's website at <http://www.nrc.gov/reading-rm/adams.html>.

Please feel free to contact Robert Gattone of my staff if you have any questions regarding this correspondence. Mr. Gattone can be reached at 630-829-9823.

Sincerely,

/RA Geoffrey M. Warren Acting for/

Aaron T. McCraw, Chief
Materials Inspection Branch
Division of Nuclear Materials Safety

Docket No. 030-02040
License No. 21-04080-01

cc: State of Michigan

EN 51341
NMED No. 150481

Arthur J. Frazier, M.D.
Radiation Safety Officer
Mount Clemens Regional Medical Center
d/b/a McLaren Macomb
1000 Harrington Blvd.
Mount Clemens, MI 48043

SUBJECT: ACKNOWLEDGMENT OF LETTER PROVIDING NOTIFICATION OF A
REPORTABLE EVENT DATED SEPTEMBER 9, 2015 – MOUNT CLEMENS
REGIONAL MEDICAL CENTER D/B/A MCLAREN MACOMB

Dear Dr. Frazier:

This acknowledges receipt of your letter dated September 9, 2015, reporting an incident that occurred on August 18, 2015, which involved an iridium-192 sealed source that got stuck in a source position check ruler during acceptance testing. The NRC has reviewed your letter and its attachments and has no further questions at this time. The incident will be examined during a future inspection.

In accordance with Title 10 of the *Code of Federal Regulations* (CFR) 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC's Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC's website at <http://www.nrc.gov/reading-rm/adams.html>.

Please feel free to contact Robert Gattone of my staff if you have any questions regarding this correspondence. Mr. Gattone can be reached at 630-829-9823.

Sincerely,

/RA Geoffrey M. Warren Acting for/

Aaron T. McCraw, Chief
Materials Inspection Branch
Division of Nuclear Materials Safety

Docket No. 030-02040
License No. 21-04080-01

cc: State of Michigan

ADAMS Accession Number: ML15264A995

DOCUMENT NAME: G:\DNMS\Work in progress\LTR - ack ltr mclaren macomb.docx

☒ Publicly Available

☐ Non-Publicly Available

☐ Sensitive

☒ Non-Sensitive

To receive a copy of this document, indicate in the concurrence box "C" = Copy without attach/encl "E" = Copy with attach/encl "N" = No copy

OFFICE	RIII DNMS		RIII DNMS		RIII		RIII	
NAME	RGattone:ps		McCraw gmw for					
DATE	09/21/15		09/21/15					

OFFICIAL RECORD COPY

Letter to Arthur M. Frazier, M.D., from Aaron T. McCraw dated September , 2015

SUBJECT: ACKNOWLEDGMENT OF LETTER PROVIDING NOTIFICATION OF A
REPORTABLE EVENT DATED SEPTEMBER 9, 2015 – MOUNT CLEMENS
REGIONAL MEDICAL CENTER D/B/A MCLAREN MACOMB

DISTRIBUTION:

Darrell Roberts
John Giessner
Julio Lara
Richard Skokowski
Carole Ariano
Paul Pelke
Carmen Olteanu
Jim Clay
MIB Inspectors