

ATTENTION: SCOTT WILSON

LICENSE NO: 31-35265-01

DOCKET NO: 03038865

CONTROL NO: 588579

SUBJECT: YOUR REQUEST FOR ADDITIONAL INFORMATION CONCERNING
APPLICATION FOR MATERIALS LICENSE CONTROL NUMBER 588579.

Dear Mr. Wilson:

This is in reference to your e-mail dated August 28, 2015, regarding my application dated August 13, 2015, requesting additional information concerning a materials license.

As per your request, I have changed application Items 8 and 10. I have also browsed websites in your e-mail and have made updates to my records.

About Item (5) in the Application Form. This Item can be replaced by "Anywhere NRC has Jurisdiction for regulating the use of radioactive material"

Mr. Wilson, I have left a phone number below where you can contact me at your convenience or preferably you can e-mail me at mohamedsharawy@aol.com.

Thank you.

Sincerely,
Mohamed Sharawy,



Phone: (212) 593-1253

On September 3rd, 2015

REC RG 109 09 15 AM 07 01

NMSS/RGNI MATERIALS-002

**ITEM (8). TRAINING FOR INDIVIUALS WORKING IN OR FREQUENTING
RESTRICTED AREA.**

- Before using licensed materials, authorized users will have successfully completed one of the training courses described in Criteria in the section entitled "Training for Individuals Working In or Frequenting Restricted Areas" in NUREG-1556, Vol. 1, Rev 1, dated November 2001."

Signed: Mohamed Sharawy



9/3/2015

ITEM (10). RADIATION SAFETY PROGRAM – SURVEY INSTRUMENTS.

We will either possess and use, or have access to and use, a radiation survey meter that meets the Criteria in the section entitled "Radiation Safety Program -Instruments" in NUREG-1556, Vol. 1, Rev. 1, dated November 2001.

SIGNED: MOHAMED SHARWAY

A handwritten signature in black ink, appearing to be 'MS' with a long horizontal stroke extending to the right.

9/3/2015

ITEM (10). RADIATION SAFETY PROGRAM – MATERIAL RECEIPT AND ACCOUNTABILITY

Physical inventories will be conducted at intervals not to exceed 6 months, to account for all sources and devices received and possessed under the license.

SIGNED: MOHAMED SHARAWY

A handwritten signature in black ink, consisting of a stylized 'S' followed by a horizontal line.

9/3/2015

ITEM (10). RADIATION SAFETY PROGRAM – OCCUPATIONAL
DOSIMETRY.

Either we will maintain, for inspection by NRC, documentation demonstrating that unmonitored individuals are not likely to receive a radiation dose in excess of 10 percent of the allowable limits in 10 CFR Part 20, or we will provide dosimetry processed and evaluated by an NVLAP-approved processor that is exchanged at a frequency recommended by the processor.

SIGNED: MOHAMED SHARAWY

A handwritten signature in black ink, consisting of a stylized 'S' followed by a horizontal line.

9/3/2015

ITEM (10), RADIATION SAFETY PROGRAM – OPERATING AND
EMERGENCY PROCEDURES.

Operating, emergency, and security procedures will be developed, implemented, and maintained and will be comparable to the Appendix H, "Operating, Emergency, and Security Procedures" dated July 5, 2005.

SIGNED: MOHAMED SHARAWY

A handwritten signature in black ink, consisting of a stylized 'M' and 'S' connected together, with a horizontal line extending to the right.

9/3/2015

ITEM (10). RADIATION SAFETY PROGRAM – LEAK TEST.

Leak tests will be performed at intervals approved by NRC or an Agreement State and specified in the Sealed Source and Device Registration Sheet. Leak tests will be performed by an organization authorized by NRC or an Agreement State to provide leak testing services for other licensees or using a leak test kit supplied by an organization authorized by NRC or an Agreement State to provide leak test kits to other licensees and according to the kit supplier's instructions.

SIGNED: MOHAMED SHARAWY


A handwritten signature in black ink, appearing to be 'MS' or similar, with a long horizontal stroke extending to the right.

9/3/2015

ITEM (10). RADIATION SAFETY PROGRAM – MAINTENANCE.

We will implement and maintain procedures for routine maintenance of our gauges according to each manufacturer's recommendations and instructions, and, We will send the gauge to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require the removal of the source or source rod from the gauge.

SIGNED: MOHAMED SHARAWY



9/3/2015