

## Walt, GERALYN

---

**From:** Reichard, Michael  
**Sent:** Wednesday, February 24, 2016 10:41 AM  
**To:** dr.cruzoftalmologo@hotmail.com  
**Cc:** crmidmr@aol.com; Lanzisera, Penny  
**Subject:** Request for additional information - Luiz M. Cruz, M.D. 52-25102-01, MC590141

Docket No. 03031733  
License No. 52-25102-01  
Control No. 590141

Luis M. Cruz, M.D.  
Dr. Basora #55 Norte  
Oficina 1 J  
Mayaguez, Puerto Rico 00680

**SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR RENEWAL TO LICENSE, CONTROL NO. 590141**

Dear Dr. Cruz:

This is in reference to your application dated January 22, 2016, requesting to renew Nuclear Regulatory Commission License No. 52-25102-01. In order to continue our review, we need the following information:

1. In reference to your facility diagram please provide the following:
  - a. Please commit that the source will be stored in its storage/shipping container.
  - b. Please provide additional information regarding security arrangements for the source. Please indicate whether the room is supervised and what doors have restricted access.
2. Please indicate how the source is verified to be secure immediately following a procedure (e.g. the source is verified with a survey instrument or by visual verification that the source is still attached to the applicator).
3. Please provide a description of the instrumentation that will be used to perform the required surveys.
4. Please ensure that your response is signed by a senior manager of the company.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits**, see our **toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 590141. If you have any technical questions regarding this deficiency letter, please call Michael Reichard at (610) 337-3645.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Michael Reichard  
Health Physicist  
Division of Nuclear Materials Safety  
United States Nuclear Regulatory Commission  
[Michael.reichard@nrc.gov](mailto:Michael.reichard@nrc.gov)

610-337-6945 (o)

610-420-5338 (c)