

**From:** [Reichard, Michael](#)  
**To:** [James.Ricci@hhchealth.org](mailto:James.Ricci@hhchealth.org)  
**Cc:** [Weidner, Tara](#)  
**Subject:** FW: Draft RAI letter  
**Date:** Tuesday, March 15, 2016 7:36:00 AM

---

Docket No. 03008728  
License No. 06-15203-01  
Control No. 590080

James Ricci  
Windham Community Memorial Hospital  
112 Mansfield Avenue  
Willimantic, Connecticut 06226

**SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING  
APPLICATION FOR  
RENEWAL TO LICENSE, CONTROL NO. 590080**

Dear Mr. Ricci:

This is in reference to your application dated January 25, 2016, requesting to renew Nuclear Regulatory Commission License No. 06-15203-01. In order to continue our review, we need the following information:

1. In reference to your facility diagram, indicate areas that are restricted vs. unrestricted and provide additional information regarding security arrangements for the byproduct material, specifically the hot lab.
2. Please verify that your use of 35.300 materials is limited to outpatients only.
3. Please provide a description of the instrumentation (e.g. gamma counter, solid state detector, portable or stationary count rate meter, portable or stationary dose rate or exposure rate meter, single or multichannel analyzer, liquid scintillation counter, proportional counter) that will be used to perform required surveys.
4. Please ensure that your response is signed by senior management.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 590080. If you have any technical questions regarding this deficiency letter, please call Michael Reichard at (610) 337-3645.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Michael Reichard

Health Physicist  
Division of Nuclear Materials Safety  
United States Nuclear Regulatory Commission  
[Michael.reichard@nrc.gov](mailto:Michael.reichard@nrc.gov)  
610-337-6945 (o)  
610-420-5338 (c)