

From: [Reichard, Michael](#)
To: melinamcastillo@gmail.com
Cc: crmidmr@aol.com; [Elliott, Robin](#)
Subject: Request for Additional Information - MC590089, 52-11897-01, 03003529
Date: Thursday, March 17, 2016 10:21:00 AM

Docket No. 03003529
License No. 52-11897-01
Control No. 590089

Roberto Buxeda Dacri, M.D.
1452 Ashford Ave.
Condominium Adalgia, Suite 409-B
San Juan, Puerto Rico 00907

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING
APPLICATION FOR
RENEWAL TO LICENSE, CONTROL NO. 590089

Dear Dr. Buxeda Dacri:

This is in reference to your application dated January 20, 2016, requesting to renew Nuclear Regulatory Commission License No. 52-11897-01. In order to continue our review, we need the following information:

1. In reference to your facility diagram, please update the diagram to include doors and access points. Provide additional information regarding security arrangements for the source. Please indicate whether the room is supervised and what doors have restricted access.
2. Please indicate how the source is verified to be secure immediately following a procedure (e.g. the source is verified with a survey instrument or by visual verification that the source is still attached to the applicator).
3. Please make the following commitment, "We have developed and will implement and maintain written procedures for area surveys in accordance with 10 CFR 20.1101 that meet the requirements of 10 CFR 20.1501 and 10 CFR 35.70." For example, surveys following use of the applicator (see 10 CFR 35.404), six month leak tests, and radiation level surveys of the storage area should be performed and your procedures should address these surveys.
4. You have indicated that you will not have a survey instrument, but a survey instrument will be available. Please indicate who will make the instrument available, and how readily available it will be.
5. Please clarify if the second line in your address is "Condominium Ada Ligia, Suite 409-B" or "Condominium Adalgia, Suite 409-B."
6. Please ensure that your response is signed by senior management.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 590089. If you have any technical questions regarding this deficiency letter, please call Michael Reichard at (610) 337-3645.

To ensure a prompt review, please submit your response in 30 calendar days.

Michael Reichard
Health Physicist
Division of Nuclear Materials Safety
United States Nuclear Regulatory Commission
Michael.reichard@nrc.gov
610-337-6945 (o)
610-420-5338 (c)