



Region I Office  
Division of Nuclear Materials Safety  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713  
(610)337-5000

## Telephone Conversation Record

Date: 4/20/2016

License No. 10-06772-01

Docket No.(no hyphens): 03004001

Mail Control/Report No. 588982

Licensee Name: U.S. Department of Health & Human Services

Participant(s) Name/Title: Narvaez L. Stinson/ Asistant Radiation Safety Officer

Work Telephone No. 404-639-2486

Business Cellphone No. 404-664-7100

NRC Representative Name/Title: Dennis Lawyer/ Health Physicist

Subject: RAI, Efficiency on Alpha Scans

Discussion: I discussed with Mr. Stinson that item 2 of my April 12, 2016, electronic mail didn't appear to be resolved. The efficiency in the calculation did not use the source efficiency of 0.25 in the calculation. After discussing with the vendor, the vendor will change the calculation and lower the scan speed.

Action Required: Change the scan speed of the alpha scan rate and resubmit the calculation.

### SUNSI REVIEW

Document Availability: ☒ Public or ☐ Non-Public

#### Document Sensitivity:

(select "1" value to the right)

<input checked="" type="checkbox"/>	Non-Sensitive	<input type="checkbox"/>	MD 3.4 Non-Public B.1 (Non-Sensitive)
<input type="checkbox"/>	Non-Sensitive Copyright	<input type="checkbox"/>	MD 3.4 Non-Public A.3 (Sensitive Security Related)
<input type="checkbox"/>	Sensitive – Proprietary	<input type="checkbox"/>	MD 3.4 Non-Public A.4 (Sensitive Proprietary)
<input type="checkbox"/>	Sensitive – Privacy Act (includes Personally Identifiable Information (PII))	<input type="checkbox"/>	MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
<input type="checkbox"/>	Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)	<input type="checkbox"/>	MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
<input type="checkbox"/>	Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/>	MD 3.4 Non-Public A.7 ( Internal)

SUNSI Review Completed by: Dennis Lawyer