



MARIE PUCKETT, CSP
Safety Manager

GENERAL MILLS

May 2, 2016

Director, Office of Nuclear Material Safety and Safeguards
ATTN: GLTS
U.S. Nuclear Regulatory Commission
Washington, DC 20555-001

Re: Transfer of Ownership (Disposal Transfer)

GL-705013-19

(NRC Materials License 13-1883-01)

General Mills
707 Pillsbury Lane
New Albany, IN 47150

Dear Sir/ Madam:

This letter serves as notification that, General Mills (license holder), has disposed of eleven (10) nuclear gauges listed below to RAM Services, Inc (Two Rivers WI) on 4/27/16.

Serial#	mCi	Wipe	Status	Housing	Initial Leak Test	Manufacturer
23664	5	Apr-16	To RAM 4/27/16	7062BP	9/1/1987	Kay-Ray
23666	5	Apr-16	To RAM 4/27/16	7062BP	9/1/1987	Kay-Ray
28360A	5	Apr-16	To RAM 4/27/16	7062BP	6/16/1990	Kay-Ray
B108	5	Feb-16	To RAM 4/27/16	7062BP	9/12/2001	Kay-Ray
B109	5	Feb-16	To RAM 4/27/16	7062BP	9/12/2001	Kay-Ray
B110	5	Apr-16	To RAM 4/27/16	7062BP	9/12/2001	Kay-Ray
B2882	10	Apr-16	To RAM 4/27/16	7062BP	3/31/1994	Kay-Ray
B52	10	Feb-16	To RAM 4/27/16	7062BP	11/8/2000	Kay-Ray
B53	10	Feb-16	To RAM 4/27/16	7062BP	11/8/2000	Kay-Ray
S96A2301	5	Feb-16	To RAM 4/27/16	7062BP	1/8/1993	Kay-Ray

The licensee contact person/ site RSO is Marie Puckett at the contact information below.

We are decommissioning this facility over the course of the next five (5) months. The plan is as follows:

- As equipment is taken out of service, any identified K-Ray associated with this equipment will be removed and placed into storage. Our specific device license allows us to do this. We will survey the device and equipment area before and after its removal from the equipment.
- The storage cabinet has always been surveyed on a regular basis and will continue to do so. Once the radiation levels for the stored devices reaches close but not over 5mR/hr at the closed cabinet door, we will call RAM Services, Inc to arrange for pick up.
- We conduct leak tests and surveys right before they are picked up by RAM to ensure there are no leaks.

May 2, 2016

- After all devices are removed I will alert your office and the Region III office of the termination of the license.
- I have attached all the relevant paperwork to this letter (RAM's receipt of materials letter; shipping documents and leak test certificates.)

Please contact me if you have any questions regarding this information.

Sincerely,

A handwritten signature in black ink that reads "Marie Puckett". The signature is fluid and cursive, with a long horizontal stroke extending from the end of the name.

Marie Puckett, CSP
Safety Manager
Radiation Safety Officer



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		137614	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NA	DEVICE/TOOL NUMBER	NA
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	23664	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	21 Apr 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	22 Apr 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	12 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	25 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-11.9053 dpm	MINIMUM DETECTABLE ACTIVITY	1.24E-05 µCi 0.46 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.22807 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato <i>Virginia Parlato</i>		
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST JOY DISCHARGE HIGH LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		137615	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NA	DEVICE/TOOL NUMBER	NA
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	23666	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	21 Apr 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	22 Apr 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	22 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	25 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-3.03158 dpm	MINIMUM DETECTABLE ACTIVITY	1.24E-05 µCi 0.46 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.05263 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST BLENDER #2 LOS LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		137616	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NA	DEVICE/TOOL NUMBER	NA
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	28360A	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	21 Apr 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	22 Apr 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	21 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	25 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-3.74737 dpm	MINIMUM DETECTABLE ACTIVITY	1.24E-05 µCi 0.46 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.07018 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato <i>Virginia Parlato</i>		
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST CO2 DUST COLLECTOR HIGH LVL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		137617	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NA	DEVICE/TOOL NUMBER	NA
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	B110	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	21 Apr 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	22 Apr 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	17 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	25 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-7.57895 dpm	MINIMUM DETECTABLE ACTIVITY	1.24E-05 µCi 0.46 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.14035 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato <i>Virginia Parlato</i>		
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST BLENDER #1 LOW LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		137618	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.	NA	DEVICE/TOOL NUMBER	NA
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	B2882	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	10 mCi 0.37 GBq
DATE SOURCE WIPED	21 Apr 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	22 Apr 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	19 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	25 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-5.74737 dpm	MINIMUM DETECTABLE ACTIVITY	1.24E-05 µCi 0.46 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.10526 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato <i>Virginia Parlato</i>	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST DUSTING FLOUR HIGH & LOW LVL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135882	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	B108	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	31 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	23 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	0.989474 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi 0.01754 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Pie Crust Blender #1 High Level			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135883	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	B109	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	31 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	19 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-2.84211 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.05263 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Pie Crust Joy Infeed High Level			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135890	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.		B52	
ISOTOPE		Cs-137	
DATE SOURCE WIPED		31 Jan 2016	
ANALYSIS DATE		04 Feb 2016	
GROSS SAMPLE COUNT		20 cpm	
BACKGROUND		22 cpm	
NET SAMPLE COUNT		-1.91579 dpm	
SAMPLE ANALYSIS ACTIVITY		0 μ CI -0.03509 Bq	
SOURCE PASSES TEST		PASS/FAIL CRITERION	
		ACTION LEVEL	
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST SHTG TANK HI LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135891	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	B53	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	10 mCi 0.37 GBq
DATE SOURCE WIPED	31 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	17 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-4.52632 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.08772 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PIE CRUST SHTG TANK LO LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135895	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	S96A2301	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	31 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	11 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-10.0737 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.19298 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: Pie Crust Blender #2 High Level			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

Notification of Disposal
Date: April 15, 2016

Waste from your organization has been processed and disposed at the following site:

**Waste Control Specialists LLC
Compact Waste Facility
9998 W State HWY 176
Andrews, TX 79714**

Attached are the certificate of disposal from Waste Control Specialists and the signed cover page of the waste manifest.

Listed below are the activity totals (in millicuries and Megabecquerels), volume (in cubic feet and cubic meters) and weight (in pounds) for your portion of the shipment.

Thank you for using RAM SERVICES INC to meet your disposal needs. Please contact us if you have any questions regarding this shipment.

Jerry P. Wiza

Jerry Wiza

RAM SERVICES INC
510 County Highway V
Two Rivers, WI 54241
www.ramservicesinc.com
920-686-3889



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

FORM 542			RAM SERVICES, INC.		2. MANIFEST NUMBER			
UNIFORM LOW-LEVEL RADIOACTIVE					2016-1100			
WASTE MANIFEST								
MANIFEST INDEX AND REGION COMPACT TABULATION								
5. GENERATOR NAME AND TELEPHONE NUMBER		6. GENERATOR FACILITY ADDRESS	8. MANIFEST NUMBER(S) UNDER WHICH WASTE (OR MATERIAL) RECEIVED AND DATE OF RECEIPT	C. ACTIVITY (MBq) (mCi)		D. VOLUME (m3) (ft3)		E. WEIGHT (pounds)
General Mills ☐ 812 941-4293		707 Pillsbury Lane ☐ New Albany, IN 47150	RPK-1037 ☐ received : 9/11/2015	7329.70	198.10	0.006	0.22	51.82
			TOTALS	7329.70	198.10	0.006	0.22	51.82



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

Notification of Disposal
Date: April 15, 2016

Waste from your organization has been processed and disposed at the following site:

**Waste Control Specialists LLC
Compact Waste Facility
9998 W State HWY 176
Andrews, TX 79714**

Attached are the certificate of disposal from Waste Control Specialists and the signed cover page of the waste manifest.

Listed below are the activity totals (in millicuries and Megabecquerels), volume (in cubic feet and cubic meters) and weight (in pounds) for your portion of the shipment.

Thank you for using RAM SERVICES INC to meet your disposal needs. Please contact us if you have any questions regarding this shipment.

Jerry P. Wiza

Jerry Wiza

RAM SERVICES INC
510 County Highway V
Two Rivers, WI 54241
www.ramservicesinc.com
920-686-3889



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

FORM 542		RAM SERVICES, INC.		2. MANIFEST NUMBER			
UNIFORM LOW-LEVEL RADIOACTIVE				2016-1100			
WASTE MANIFEST							
MANIFEST INDEX AND REGION COMPACT TABULATION							
5. GENERATOR NAME AND TELEPHONE NUMBER	6. GENERATOR FACILITY ADDRESS	8. MANIFEST NUMBER(S) UNDER WHICH WASTE (OR MATERIAL) RECEIVED AND DATE OF RECEIPT	C. ACTIVITY (MBq) (mCi)		D. VOLUME (m3) (ft3)	E. WEIGHT (pounds)	
General Mills☐ 812-941-4293	707 Pillsbury Lane☐ New Albany, IN 47150	RPK-1088☐ received : 11/21/2015	2361.71	63.83	0.006	0.22	51.82
		TOTALS	2361.71	63.83	0.006	0.22	51.82