

License: 25-27721-01
Docket: 030-35944
Control: 590210

Cook, Jackie

To: Kari Cann
Cc: angie.korte@gfclinic.com
Subject: RE: Re: Amendment Request dated February 10, 2016

Thank you Kari for your prompt response. I appreciate you trying to get your response to me by Thursday, May 12, 2016. I will mark my calendar accordingly.

In response to your question below on how long a hospital need to retain paperwork after releasing a facility for unrestricted release: Please note that in accordance with 10 CFR 30.35(g) which states in part that each person licensed under this part or parts 32 through 36 and 39 of this chapter shall keep records of information important to the decommissioning of a facility in an identified location until the site is released for unrestricted use. In addition, please note that the hospital may want to keep those records for some other reasons. As you know we keep up with all your documents electronically through ADAMS and once you throw away hard copies of your records but later want to retrieve those records for some reason, you can submit a FOIA request and we can retrieve the records from ADAMS for you.

Please do not hesitate to contact me if I can be of further assistance.

Warmest Regards,

Ms. Jacqueline "Jackie" D. Cook
Senior Health Physicist
US Nuclear Regulatory Commission Region IV
Division of Nuclear Materials Safety
Nuclear Materials Safety Branch B
1600 East Lamar Blvd., Arlington, TX 76011
817-200-1132 (office)/817-200-1263 (fax)
Email address: Jackie.Cook@nrc.gov

From: Kari Cann [mailto:kndcann@aol.com]
Sent: Thursday, May 05, 2016 3:16 PM
To: Cook, Jackie <Jackie.Cook@nrc.gov>
Cc: angie.korte@gfclinic.com; KNDCANN@aol.com
Subject: [External_Sender] Re: Amendment Request dated February 10, 2016

Hello Jackie

Great to hear from you! It really has been a while since we communicated on a regular basis

We will look again for the final status surveys for the 9th Street location. We have had significant employee turnover and disgruntled ex employees so we did not have any success when we looked before. We may need to re survey the area and if that is the case, I will send you the current survey and wipes. I was hoping the statement signed by me that we originally submitted would suffice The only material in the room was Tc 99 (with a 6 hour half life) and that was 6 months ago. I can however see the need for paperwork completeness.

One more question: I have a different site that closed last year and Mr. Torres approved the closure and the submitted surveys and wipe tests. The hospital is now asking how long they need to retain the paperwork showing they were a licensed facility. I'm not finding the right location on the website to answer that.

PUBLIC

☐ Immediate Release

☒ ~~CONFIDENTIAL~~

NON-PUBLIC

☐ A.3 Sensitive-Security Related

☐ A.7 Sensitive Internal

☐ Other:

Reviewer: JAC Date: 5/5/16

You will receive something from me by May 12 regarding the Great Falls Clinics 9th Street location final surveys

As Always
Kari

Kari Cann, MS DABR
Medical Physicist and RSO
Great Falls Clinic Medical Center
Great Falls, MT
406-788-7887

Good Afternoon Kari: How are you? Long time no talk to. I hope you've been doing well.

Regarding your amendment request dated February 10, 2016, requesting to release 2 locations of use for unrestricted use, after reading your reason for why the location on 9th Street was ever specified on the license, you do state that occasionally patients with Tc-99m were admitted to the surgical suite. Please submit copies of the final status surveys (specific wipe report/exit surveys) as you did for the other location on 15th Avenue South.

Please submit the final status survey for the location on 9th Street by May 12, 2016.

Thanking you in advance for your prompt response. Please do not hesitate to contact me at the phone numbers or email address below if you have any questions regarding the above.

Warmest Regards,

Ms. Jacqueline "Jackie" D. Cook
Senior Health Physicist
US Nuclear Regulatory Commission Region IV
Division of Nuclear Materials Safety
Nuclear Materials Safety Branch B
1600 East Lamar Blvd., Arlington, TX 76011
817-200-1132 (office)/817-200-1263 (fax)
Email address: Jackie.Cook@nrc.gov

-----Original Message-----

From: Cook, Jackie <Jackie.Cook@nrc.gov>

To: kndcann <kndcann@aol.com>

Sent: Wed, May 4, 2016 4:24 pm

Subject: Amendment Request dated February 10, 2016

Good Afternoon Kari: How are you? Long time no talk to. I hope you've been doing well.

Regarding your amendment request dated February 10, 2016, requesting to release 2 locations of use for unrestricted use, after reading your reason for why the location on 9th Street was ever specified on the license, you do state that occasionally patients with Tc-99m were admitted to the surgical suite. Please submit copies of the final status surveys (specific wipe report/exit surveys) as you did for the other location on 15th Avenue South.

Please submit the final status survey for the location on 9th Street by May 12, 2016.

Thanking you in advance for your prompt response. Please do not hesitate to contact me at the phone numbers or email address below if you have any questions regarding the above.

Warmest Regards,

Ms. Jacqueline "Jackie" D. Cook

Senior Health Physicist

US Nuclear Regulatory Commission Region IV

Division of Nuclear Materials Safety

Nuclear Materials Safety Branch B

1600 East Lamar Blvd., Arlington, TX 76011

817-200-1132 (office)/817-200-1263 (fax)

Email address: Jackie.Cook@nrc.gov