

UNITED STATES
NUCLEAR REGULATORY COMMISSION

REGION III
2443 WARRENVILLE RD. SUITE 210
LISLE, IL 60532-4352

JUN 02 2016

James M. Botti, MS
Radiation Safety Officer
National Diagnostic Service, LLC
39595 W. Ten Mile Rd.
Suite 102
Novi, MI 48375

Dear Mr. Botti:

Enclosed is Amendment No. 09 terminating your NRC Materials License No. 21-32432-01 in accordance with your request.

In letter dated March 30, 2016, you contacted the U. S. Nuclear Regulatory Commission and indicated that you wished to terminate NRC Materials License No. 21-32432-01 for National Diagnostic Service, LLC. The aforementioned letter included NRC Form 314, "Certificate of Disposition of Materials," dated March 29, 2016, with attachments, supporting the decommissioning of your mobile van unit based at 39595 West Ten Mile Road, Suite 102, Novi, Michigan, and your location of use at 39475 Lewis Drive, Suite 200, Novi, Michigan. Additional supporting information was provided by you and Ms. Jennifer McEvers upon our request.


The NRC staff has reviewed all of this information, including your final status surveys. Based on its review, the staff has concluded that all licensable radioactive material has been removed from your facility and residual radioactive material attributable to licensed activities does not exceed current NRC criteria.

Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required. Your mobile van unit based at 39595 West Ten Mile Road, Suite 102, Novi, Michigan, and your facility at 39475 Lewis Drive, Suite 200, Novi Michigan, are both suitable for unrestricted use and NRC License No. 21-32432-01 is hereby terminated.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U. S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

In accordance with Title 10 of the *Code of Federal Regulations* Section 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Sincerely,

A handwritten signature in black ink, appearing to read "BAPh", is written over the typed name.

Bryan A. Parker
Senior Health Physicist
Materials Licensing Branch

License No. 21-32432-01
Docket No. 030-36165

Enclosure: Amendment No. 09 (Termination)