

U.S. NUCLEAR REGULATORY COMMISSION**Date:** May 27, 2016**TELEPHONE CONVERSATION RECORD**

Mail Control No.	Not assigned yet	License No.	06-06697-02	Docket No.	030-01265
-------------------------	-------------------------	--------------------	-------------	-------------------	-----------

Name of Licensee: Stamford Hospital

Name of Participant(s): Sarah Bull, RSO
Shawn Seeley, Health Physicist, RI DNMS

Telephone No. 203-276-4036

Subject: Questions to Amendment Application

(NOTE: This will be used as the Documents Title in ADAMS)

Ms. Bull called and left a message on the division mailbox with questions regarding an amendment letter that she recently submitted for changes to the HDR program. I returned her call and found out they were changing the room where the HDR was to be used and had run into a construction delay.

She asked whether it was possible to decommission the HDR until such time as the renovations to the room were completed. I told her that was an acceptable solution. I reminded her that in any case, the HDR unit should be secured from unauthorized access while not in use during renovations to the area/room.

I told her to call back if she had any further questions.

Document Availability:☒**Publicly Available**☐**Non-Publicly Available**☒**Non-Sensitive**☐**Non-Sensitive Copyright**☐**Sensitive- Proprietary****Sensitive – Privacy Act/ PII**☐**Sensitive – Internal****Sensitive – Security-Related****Immediate Release****Normal Release Date:****Delayed Release Date:****SUNSI Review Completed By:** SSeeley / RA /**Document Accession #:** _____