

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

## 1. LICENSEE/LOCATION INSPECTED:

Heart Care  
6889 Highland Road  
Waterford, Michigan 48327

REPORT NUMBER(S) 2016001

## 2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

## 3. DOCKET NUMBER(S)

030-30281

## 4. LICENSE NUMBER(S)

21-25851-01

## 5. DATE(S) OF INSPECTION

May 23, 2016

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

## Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Navid Tehrani		5/23/16
BRANCH CHIEF	Aaron McCraw		6/7/16

**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:

Heart Care  
6889 Highland Road  
Waterford, Michigan 48327

REPORT NUMBER(S) 2016001

2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-30281

4. LICENSE NUMBER(S)

21-25851-01

5. DATE(S) OF INSPECTION

May 23, 2016

6. INSPECTION PROCEDURES USED

87130

7. INSPECTION FOCUS AREAS

03.01-03.08

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)

02201

2. PRIORITY

5

3. LICENSEE CONTACT

Creagh E. Milford D.O., RSO

4. TELEPHONE NUMBER

(248) 666-5200



Main Office Inspection

Next Inspection Date:

05/23/2021



Field Office Inspection



Temporary Job Site Inspection

**PROGRAM SCOPE**

This was an unannounced routine inspection of a private practice authorized by its NRC license to use licensed material authorized by 10 CFR 35.100 and 35.200 at its facility in Waterford, MI. The licensee receives unit doses from a local radiopharmacy for a wide variety of diagnostic procedures (including MUGA). The hospital employs two part-time technologists who perform approximately 15 to 25 diagnostic procedures per month, Monday through Friday. The clinic retains the services of a medical physics consultant to perform instrument calibrations, sealed source leak tests, quarterly audits of the radiation safety program and annual ALARA audits.

Interviews of available staff revealed an adequate level of understanding of emergency and material handling procedures and techniques. Dose calibrator constancy checks, package receipt, daily surveys, and waste handling and disposal procedures were successfully demonstrated. The inspector was also able to observe the preparation and administration of two resting cardiac stress doses. An outside consultant performed quarterly program audits that were adequate to oversee the program. Licensed material was adequately secured and not readily accessible to members of the general public. The licensee possessed a radiation survey meter that was calibrated, operational, and performed well in side-by-side comparison with an NRC instrument. Independent measurements did not indicate readings in excess of Title 10 of the Code of Federal Regulations (10 CFR) Part 20 limits in restricted or unrestricted areas. Personal whole body and extremity dosimetry were observed worn by the staff during the inspection, and records did not indicate doses in excess of 10 CFR Part 20 limits.

No violations of NRC requirements were identified during this inspection.