



MARIE PUCKETT, CSP
Safety Manager

GENERAL MILLS

June 13, 2016

Director, Office of Nuclear Material Safety and Safeguards
ATTN: GLTS
U.S. Nuclear Regulatory Commission
Washington, DC 20555-001

Re: Transfer of Ownership (Disposal Transfer)

GL-705013-19

(NRC Materials License 13-1883-01)

General Mills

707 Pillsbury Lane

New Albany, IN 47150

Dear Sir/ Madam:

This letter serves as notification that, General Mills (license holder), has disposed of twelve (12) nuclear gauges listed below to RAM Services, Inc (Two Rivers WI) on 6/13/16.

Manufacturer	Serial#	Initial Leak Test	Most Recent Leak Test
Kay-Ray	18838-C	3/5/1985	2/4/2016
Kay-Ray	18838-D	3/5/1985	2/4/2016
Kay-Ray	18838-E	3/5/1985	2/4/2016
Kay-Ray	18838-G	3/5/1985	2/4/2016
Kay-Ray	18838-H	3/5/1985	2/4/2016
Kay-Ray	18838-J	3/5/1985	2/4/2016
Kay-Ray	18838-K	3/5/1985	2/4/2016
Kay-Ray	18838-L	3/5/1985	2/4/2016
Kay-Ray	18838-M	3/5/1985	2/4/2016
Kay-Ray	18838-N	3/5/1985	2/4/2016
Kay-Ray	18838-O	3/5/1985	2/4/2016
Kay-Ray	20398	4/2/1986	2/4/2016

The licensee contact person/ site RSO is Marie Puckett at the contact information below.

We are decommissioning this facility over the course of the next three (3) months. The plan is as follows:

- As equipment is taken out of service, any identified K-Ray associated with this equipment will be removed and placed into storage. Our specific device license allows us to do this. We will survey the device and equipment area before and after its removal from the equipment.

June 13, 2016

- The storage cabinet has always been surveyed on a regular basis and will continue to do so. Once the radiation levels for the stored devices reaches close but not over 5mR/hr at the closed cabinet door, we will call RAM Services, Inc to arrange for pick up.
- We conduct leak tests and surveys right before they are picked up by RAM to ensure there are no leaks.
- After all devices are removed I will alert your office and the Region III office of the termination of the license.
- I have attached all the relevant paperwork to this letter (RAM's receipt of materials letter; shipping documents and leak test certificates.)

Please contact me if you have any questions regarding this information.

Sincerely,

A handwritten signature in black ink, reading "Marie Puckett". The signature is fluid and cursive, with a long horizontal stroke extending from the end of the name.

Marie Puckett, CSP
Safety Manager
Radiation Safety Officer



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE	
RAM ACCESSION NO.	135850
GENERAL MILLS - NEW ALBANY	
LOCATION	47150 - INDIANA
CONTACT PERSON	MARIE PUCKETT
PHONE	812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America	
LEAK TEST ANALYSIS	
EQUIPMENT NO.	DEVICE/TOOL NUMBER
MODEL #	7062BP
SOURCE SERIAL NO.	18838-C
ISOTOPE	Cs-137
DATE SOURCE WIPED	30 Jan 2016
ANALYSIS DATE	04 Feb 2016
GROSS SAMPLE COUNT	12 cpm
BACKGROUND	22 cpm
NET SAMPLE COUNT	-10 dpm
SAMPLE ANALYSIS ACTIVITY	0 μ Cl -0.17544 Bq
SOURCE PASSES TEST	
ANALYST	Virginia Parlato
NOTES:	
ADDITIONAL SOURCE DESCRIPTION: PHL SALT HI LEVEL	



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135851	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-D	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	15 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-6.77895 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.12281 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL E-SODA HI LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135852	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-E	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	13 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-8.62105 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.15789 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL GLUTEN LO LEVEL, A2			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135854	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-G	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	21 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-0.957895 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.01754 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: CFL GLUTEN LO LEVEL, B1			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135855	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-H	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	20 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-1.95789 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.03509 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: CFL GLUTEN LO LEVEL, B2			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135857	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-J	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	13 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-7.95789 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.15789 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL E-SODA LO LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135858	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-K	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	18 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-3.78947 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.07018 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL GDL LO LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135859	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-L	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	16 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-5.81053 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.10526 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL SALT LO LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135860	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-M	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	15 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-6.77895 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.12281 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL GDL HI LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135861	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-N	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	11 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-9.61053 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.19298 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL SUGAR HI LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135862	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	18838-O	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	4 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-15.3474 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	-1E-05 µCi -0.31579 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST		Virginia Parlato	
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL SUGAR LO LEVEL			



510 County Highway V
Two Rivers, WI 54241
Voice +1-920-686-3889
Fax +1-920-686-3899

STATE OF WISCONSIN RADIOACTIVE MATERIALS LICENSE 071-1234-01

SEALED SOURCE LEAK TEST CERTIFICATE			
RAM ACCESSION NO.		135866	
GENERAL MILLS - NEW ALBANY			
LOCATION		47150 - INDIANA	
CONTACT PERSON		MARIE PUCKETT	PHONE 812-941-4293
GENERAL MILLS NEW ALBANY 707 PILLSURY LANE NEW ALBANY, INDIANA 47150 United States of America			
LEAK TEST ANALYSIS			
EQUIPMENT NO.		DEVICE/TOOL NUMBER	
MODEL #	7062BP	MANUFACTURER	Kay-Ray
SOURCE SERIAL NO.	20398	MANUFACTURER S/N	
ISOTOPE	Cs-137	SOURCE ORIGINAL ACTIVITY	5 mCi 0.185 GBq
DATE SOURCE WIPED	30 Jan 2016	REASON FOR TEST	PERIODIC
ANALYSIS DATE	04 Feb 2016	ANALYTICAL METHOD	Liquid Scintillation Counting
GROSS SAMPLE COUNT	21 cpm	INSTRUMENT	Perkin Elmer Tri-Carb 3110TR 130414
BACKGROUND	22 cpm	COUNTING EFFICIENCY	95%
NET SAMPLE COUNT	-0.989474 dpm	MINIMUM DETECTABLE ACTIVITY	1.18E-05 µCi 0.44 Bq
SAMPLE ANALYSIS ACTIVITY	0 µCi -0.01754 Bq	PASS/FAIL CRITERION	0.005 µCi 185 Bq
SOURCE PASSES TEST		ACTION LEVEL	0.001 µCi 37 Bq
ANALYST	Virginia Parlato <i>Virginia Parlato</i>		
NOTES:			
ADDITIONAL SOURCE DESCRIPTION: PHL Gluten One Batch Level - A2			

FORM 540 UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST SHIPPING PAPER		RAM SERVICES, INC 5. SHIPPER - NAME AND FACILITY General Mills 707 Pillsbury Lane New Albany IN 47160		SHIPPER I.D. NUMBER 47160 COLLECTOR PROCESSOR GENERATOR TYPE IND		7. NRC FORM 540 AND 540A PAGE 1 OF 1 PAGE(S) NRC FORM 541 AND 541A 1 PAGE(S) NRC FORM 542 AND 542A 1 PAGE(S) ADDITIONAL INFORMATION 1 PAGE(S)		8. MANIFEST NUMBER (Use this number on all continuation pages) 2016-0063	
1. EMERGENCY TELEPHONE NUMBER (Include Area Code) 800-424-9300		USER PERMIT NUMBER 2016-0063		SHIPMENT NUMBER 2016-0063		9. CONSIGNEE - Name and Facility Address RAM Services, Inc. 510 COUNTY HIGHWAY V TWO RIVERS WI 54241		CONTACT JERRY WIZA, RPT TELEPHONE NUMBER (Include Area Code) 1-920-686-3889	
ORGANIZATION CHEMTREC [for RAM SERVICES CONTRACT 210864]		CONTACT Marie Puckett TELEPHONE NUMBER 812 941-4293		9. CARRIER - Name and Address RAM Services, Inc. 510 COUNTY HIGHWAY V TWO RIVERS WI 54241		EPA I.D. NUMBER NA SHIPPING DATE 16-Jun-2016		10. CERTIFICATION This is to certify that the herein-named materials are properly classified, described, packaged, marked, and labeled and are in proper condition for transportation according to the applicable regulations of the Department of Transportation. This also certifies that the materials are classified, packaged, marked, and labeled and are in proper condition for transportation and disposal as described in accordance with the applicable requirements of 10 CFR Parts 20 and 61, or equivalent state regulations.	
2. IS THIS AN "EXCLUSIVE USE" SHIPMENT? <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		3. TOTAL NUMBER OF PACKAGES IDENTIFIED ON THIS MANIFEST 1		EPA MANIFEST NUMBER NOT APPLICABLE		CONTACT JERRY WIZA, RPT SIGNATURE - Authorized carrier acknowledging waste receipt <i>[Signature]</i>		DATE 13 JUNE 16	
4. DOES EPA REGULATED WASTE REQUIRING A MANIFEST ACCOMPANY THIS SHIPMENT? <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO If "Yes," provide Manifest Number:		11. U.S. DEPARTMENT OF TRANSPORTATION DESCRIPTION (Including UN ID number, proper shipping name, hazard class, and any additional information)		12. DOT LABEL "RADIOACTIVE"		13. TRANSPORT INDEX		14. PHYSICAL AND CHEMICAL FORM	
UN2915, Radioactive material, type A package, 7		Yellow II		0.3		SOLID OXIDE		Cs-137 INDIVIDUAL RADIONUCLIDES mCi MBq 2.92E+01 1.08E+03	
18. TOTAL PACKAGE ACTIVITY IN SI UNITS		17. LSA/SCO CLASS		18. TOTAL WEIGHT OR VOLUME (Use appropriate units)		19. IDENTIFICATION NUMBER OF PACKAGE		NONE 181.4 Kg 400.0 Lbs 7.5 FT3 RPK-1173	
FOR CONSIGNEE USE ONLY _____ Record Waste Description Inadequate _____ Contamination or Leakage Detected _____ Unexpected Exposure Rates Detected _____ Labels, Markings, etc. Inadequate _____ Container Integrity Inadequate _____ Other _____ No Violations Detected on this Shipment									

FORM 541		RAM SERVICES, INC.		1. MANIFEST TOTALS							2. MANIFEST NUMBER																																																																																																																																
UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST CONTAINER AND WASTE DESCRIPTION Additional Nuclear Regulatory Commission (NRC) Requirements for Control, Transfer and Disposal of Radioactive Waste										NUMBER OF PACKAGES/ DISPOSAL CONTAINERS		NET WASTE VOLUME		NET WASTE WEIGHT		SPECIAL NUCLEAR MATERIAL (grams)				2016-0063																																																																																																																							
																U-233		U-235		Pu		TOTAL		3. PAGE 1 OF 1																																																																																																																			
										1		m3 0.16 R3 5.63		kg 156.49 lb 345		NP		NP		NP		0		4. SHIPPER NAME																																																																																																																			
										ACTIVITY (MBq/mCi)										SOURCE				General Mills																																																																																																																			
										ALL NUCLIDES										TRITIUM		C-14		Tc-99		I-129		SHIPPER ID NUMBER																																																																																																															
MBq										1.08E+03		NP		NP		NP		kg NP		47150																																																																																																																							
mCi										2.92E+01		NP		NP		NP		lb NP																																																																																																																									
DISPOSAL CONTAINER DESCRIPTION										WASTE DESCRIPTION FOR EACH WASTE TYPE IN CONTAINER										16. WASTE CLASSIFICATION																																																																																																																							
5. CONTAINER IDENTIFICATION NUMBER/ GENERATOR ID NUMBER(S)										11. PHYSICAL DESCRIPTION										14. CHEMICAL DESCRIPTION										15. RADIOLOGICAL DESCRIPTION										AS-Class A AU- Class A Unstable B-Class B C-Class C																																																																																																			
6. CONTAINER DESCRIPTION (See Note 1)										12. APPROXIMATE WASTE VOLUME(S) IN CONTAINER										13. SORBENT SOLIDIFICATION STABILIZATION MEDIA (See Note 3)										17. INDIVIDUAL RADIONUCLIDES, ACTIVITY, CONCENTRATION, AND CONTAINER TOTAL ACTIVITY																																																																																																													
7. VOLUME m ³ ft ³										8. WASTE AND CONTAINER WEIGHT kg lb										9. SURFACE RADIATION LEVEL mSv/hr mrem/hr										10. SURFACE CONTAMINATION MBq/100cm ² dpm/100cm ²										18. RADIONUCLIDES																																																																																																			
RPK-1173										0.0/										<3.6E-5										<3.6E-5										36										0.16										100										SOLID OXIDE										NP										Cs-137										1.87E+05										1.08E+03										2.92E+01										AS-Class A									
/47150										4										7.5										400										1.0										<220										<2200										5.63										100										NONE																				SUBTOTAL										1.08E+03														2.92E+01					

FORM 542		RAM SERVICES, INC.		1. WASTE COLLECTOR/PROCESSOR						2. MANIFEST NUMBER					
UNIFORM LOW-LEVEL RADIOACTIVE WASTE MANIFEST MANIFEST INDEX AND REGION COMPACT TABULATION List all original "PROCESSED WASTE" generators (if any) before "COLLECTED WASTE" generators.				Name		SHIPPER USE ONLY				2016-0063					
				RAM Services, Inc.											
				IDENTIFICATION NUMBER											
				64241											
				7 Shipping Date						3					
				16-Jun-2016						Page 1 of 1					
4. GENERATOR IDENTIFICATION NUMBER	5. GENERATOR NAME AND TELEPHONE NUMBER	6. GENERATOR FACILITY ADDRESS	6A. WASTE DESCRIPTION (NOMENCLATURE)	7. PREPROCESSED WASTE (OR MATERIAL) VOLUME (m3) (R3)	8. MANIFEST NUMBER(S) UNDER WHICH WASTE (OR MATERIAL) RECEIVED AND DATE OF RECEIPT	9. WASTE CODE P= PROCESSED C= COLLECTED	10. ORIGINATING COMPACT REGION OR STATE	11. AS PROCESSED/COLLECTED TOTAL							
								A. SOURCE MATERIAL (kg) (lb)	B. SNM (g)	C. ACTIVITY (MBq) (mCi)	D. VOLUME (m3) (B3)	E. WEIGHT (pounds)	F. MAXIMUM PACKAGE RADIATION LEVEL (mrem/hr)		
47150	General Mills 812 941-4293	707 Pillsbury Lane New Albany, IN 47150	Sealed Source/Device	0.212376 7.50	RPK-1173 received : 6/16/2016	C	IN	0 0	0	1.08E+03 2.92E+01	0.212 7.50	490.00			
TOTALS OF ALL PAGES(FORMS 542 AND 542A)								0.00E+00	0.00E+00	0.00E+00	1.08E+03 2.92E+01	2.12E-01 7.50E+00	4.00E+02		