



July 29, 2016

U.S. Nuclear Regulatory Commission
2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406-2713

Br.1

Re: Termination of Radioactive Materials License No. 06-30265-01 / 03034024

REC'D 10804 15 00555

This letter is intended to notify you of the decommissioning of our Nuclear Medicine Department. As such, we are requesting a termination our Radioactive Materials License. As of June 2016, we no longer service Nuclear Medicine patients or hold radioactive materials on site.

On June 7, 2016, all long lived sealed sources were transferred to our sister company Diagnostic Imaging of Southbury, Radioactive Materials License No. 06-31024-01 via Cardinal Health in compliance with DOT regulations.

We performed area and wipe surveys ensuring all areas and surfaces were below background radiation levels.

Thank you for terminating our license. If you have any questions regarding this issue, please contact me at (203) 758-7700 or at bouellette@nvranet.com.

Sincerely,

Beth Ouellette, CNMT
Operations Manager
Naugatuck Valley Radiology

cc: Dr. Michael Malin, R.S.O

591653

NVCS/RGNI MATERIALS-002



CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

Naugatuck Valley Radiology Assoc.
385 Main St. S. Bldg 1 St. 204
Southbury Ct 06488

LICENSE NUMBER

06-30265-01

DOCKET NUMBER

030-34024

LICENSE EXPIRATION DATE

10/31/2021

A. LICENSE STATUS (Check the appropriate box)

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☒ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- ☒ a. Transfer of radioactive materials to the licensee listed below:
Diagnostic Imaging of Southbury
License # 06-31024-01
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:
- ☒ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- ☒ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☒ a. the absence of licensed radioactive materials
- ☒ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☒ 2. A copy of the radiation survey results:
- ☒ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date _____
- ☐ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☐ a. The results of the latest leak test are attached; and/or ☐ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Beth Ducikette, CNMT	Operations Manager	203-758-7700	boucllette@nuranet.com

Mail all future correspondence regarding this license to:

166 Waterbury Rd St. 105, Prospect, Ct 06712

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Beth Ducikette, CNMT
Operations Manager

SIGNATURE

[Signature]

DATE

10/29/16

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

NAUGATUCK VALLEY RADIOLOGY
166 WATERBURY RD. ST. 105
PROSPECT, CT 06712
203-758-7700

AREA SURVEY RESULTS REPORT

Test Date: 6/7/2016

Instrument: Ludlum 14-C

Serial # 12709

Background: .02 mR/hr

Test Initials: BAO

Test Location	Area Type	Reading mR/hr	Net results mR/hr	Action level mR/hr
Cardiac Left	Non-Restricted	.01	-.01	.5
Forte Camera	Non-Restricted	.02	.00	.5
Wk Station	Non-Restricted	.01	-.01	.5
Keyboard	Non-Restricted	.02	.00	.5
Uptake & Probe	Non-Restricted	.02	.00	.5
Scan Rm Trash	Non-Restricted	.03	.01	.5
Inj Rm Trash	Non-Restricted	.03	.01	.5
Inj Chair	Non-Restricted	.03	.01	.5
L Block Shield	Restricted	.07	.05	2
Sink Area	Restricted	.05	.03	2
Prep Area	Restricted	.07	.05	2
Strg Hot Trash	Restricted	.04	.02	2

Performed by: Beth Ouellette, CNMT



NAUGATUCK VALLEY RADIOLOGY
166 WATERBURY RD. ST. 105
PROSPECT, CT 06712
203-758-7700

AREA WIPE SURVEY RESULTS REPORT

Test Date: 6/7/2016

Instrument: AtomLab 950

Serial # 25738311889

Background: 0 dpm

Test Initials: BAO

Test Location	Area Type	Wipe Dpm	Net results Dpm	Action level Dpm
Strg Hot Trash	Restricted	0	0	2000
Prep Area	Restricted	0	0	2000
Sink Area	Restricted	0	0	2000
Wk Station	Non-Restricted	0	0	200
Keyboard	Non-Restricted	0	0	200
Inj Chair	Non-Restricted	0	0	200
Uptake & Probe	Non-Restricted	0	0	200
Scan Rm Trash	Non-Restricted	0	0	200
Inj Rm Trash	Non-Restricted	0	0	200
Forte Camera	Non-Restricted	0	0	200
Cardiac Left	Non-Restricted	0	0	200
Cardiac Middle	Non-Restricted	0	0	200
Cardiac Right	Non-Restricted	0	0	200

Performed by: Beth Ouellette, CNMT



NAUGATUCK VALLEY RADIOLOGY
166 WATERBURY RD. ST. 105
PROSPECT, CT 06712
203-758-7700

SEALED SOURCE WIPE RESULTS REPORT

Test Date: 6/7/2016

Instrument: AtomLab 950

Serial # 25738311889

Background: 0 dpm


Test Initials: BAO

Co57 Sheet Source 0 dpm

Cs 137 Vial 0 dpm

Ba133 Vial 0 dpm

Performed by: Beth Ouellette, CNMT

A handwritten signature in black ink, appearing to be 'BO' or 'BHO', located below the text 'Performed by: Beth Ouellette, CNMT'.

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	In accordance with the letter dated May 18, 2012,
1. Naugatuck Valley Radiological Associates	3. License number 06-30265-01 is amended in its entirety to read as follows:
2. Union Square Plaza Building 1 385 Main Street South Southbury, Connecticut 06488	4. Expiration date October 31, 2021
	5. Docket No. 030-34024 Reference No.

6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Any byproduct material permitted by 10 CFR 35.100	A. Any	A. As needed
B. Any byproduct material permitted by 10 CFR 35.200	B. Any	B. As needed
9. Authorized use:		
A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.		
B. Any imaging and localization study permitted by 10 CFR 35.200.		

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 166 Waterbury Road, Suite 105, Prospect, Connecticut.
11. The Radiation Safety Officer for this license is Michael Malin, M.D.
12. Licensed material is only authorized for use by, or under the supervision of:
- A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for medical use as indicated:

Authorized Users

Matthew Bushey, M.D.

Material and Use

35.100; 35.200

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
06-30265-01Docket or Reference Number
030-34024

Amendment No. 17

Authorized UsersMaterial and Use

Raymond Gibney, M.D.

35.100; 35.200

Robert G. Gumbardo, M.D.

35.100; 25.200

Jeffrey H. Hawley, M.D.

35.100; 35.200

Robert M. Lehman, M.D.

35.100; 35.200

Michael Malin, M.D.

35.100; 35.200

Paolo Olcese, M.D.

35.100; 35.200

Scott Segal, M.D.

35.100; 35.200

Carolyn Zajac Ajemian, M.D.

35.100; 35.200

13. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.
14. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
06-30265-01Docket or Reference Number
030-34024

Amendment No. 17

15. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated May 6, 2011 (ML111290243)
- B. Letter dated October 24, 2011 (ML112980195)
- C. Letter dated December 1, 2011 (ML113360447)
- D. Letter dated June 12, 2012 (ML12177A022)

For the U.S. Nuclear Regulatory Commission

Date July 26, 2012

By

Original signed by Maryann Abogunde

Maryann Abogunde
Medical Branch
Division of Nuclear Materials Safety
Region I
King of Prussia, Pennsylvania 19406
Thursday, July 26, 2012 11:21:41



ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee

Naugatuck Valley Radiological Associates
ATTN: Beth Ouellette, Office Manager
385 Main Street South
Union Square, Plaza Bldg., 1
Southbury, CT 06488

Date

August 9, 2016

License Number(s)

06-30265-01

Mail Control Number(s)

591653

Licensing and/or Technical Reviewer or Branch

Medical Branch (Branch 1)

This is to acknowledge receipt of your: ☒ Letter and/or ☐ Application Dated: July 29, 2016

The initial processing, which included an administrative review, has been performed.

☐ Amendment ☒ Termination ☐ New License ☐ Renewal

☒ There were no administrative omissions identified during our initial review.

☐ This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

☐ Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>
Follow the instructions on the form for submission.

☐ The following administrative omissions have been identified:

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region I
U. S. Nuclear Regulatory Commission
Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713
(610) 337-5260, (610) 337-5313,
(610) 337-5398, or (610) 337-5239