



 **BRONSON**
Methodist Hospital

August 31, 2016

U. S. NUCLEAR REGULATORY COMMISSION
Region III
Materials Licensing Branch
2443 Warrenville Road, Ste 210
Lisle, IL 60532-4352

RE: Additional information for amendment request of NRC License Control # 591665

As requested information:

- (1) The form 314 is attached.
- (2) There is no PET radionuclides to be used at the requested location.
- (3) I was present on August 31, 2016 for the quarterly physics review of the site, and there was a radiation survey performed, a recently calibrated GM EWGM was used, all surveys were appropriate. Wipe tested was performed, all areas measured background. I also reviewed recent site survey records, all were appropriate. There is nothing to report, no leaking sources.

If added information or further clarification is needed, we request that this be done via email for Laura T. Smith, MS, DABR - physicist at lsphysics@att.net and slaura@bronsonhg.org

Sincerely,



Laura T. Smith, MS, DABR
Radiation Safety Officer

601 John Street
Kalamazoo, MI 49007
269.341.7654
bronsonhealth.com



CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollections.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

BRONSON Lakeview Hospital
408 Hazen Street
Paw Paw, MI 49079-

LICENSE NUMBER

21-26716-01

DOCKET NUMBER

LICENSE EXPIRATION DATE

A. LICENSE STATUS (Check the appropriate box)

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☐ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- ☒ a. Transfer of radioactive materials to the licensee listed below:
BRONSON Methodist Hospital 21-13125-01.
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:
- ☐ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- ☒ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☒ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☒ 2. A copy of the radiation survey results:
- ☒ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date _____
- ☐ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☐ a. The results of the latest leak test are attached; and/or ☐ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Laura T. Smith	RSD	566 808 3058	lschysics@att.net slaura@bronsonhwy.org

Mail all future correspondence regarding this license to:

same location - still open & just merging to another license.

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

SIGNATURE

DATE

Laura T. Smith, RSD

Laura T. Smith

8/31/2016

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

Forster, Sara

From: Laura T. Smith- Physics <lsphysics@att.net>
Sent: Wednesday, August 31, 2016 11:09 AM
To: Forster, Sara
Subject: [External_Sender] Re: Additional Information request for , CN591664 & CN59166
Attachments: SCAN0323.PDF; SCAN0324.PDF

Sara,
This should complete what I have outstanding for you. Any questions, or if something is not as requested please let me know asap, and I am back now, so will get done quickly.

Laura Smith I prefer email communications

From: "Forster, Sara" <Sara.Forster@nrc.gov>
To: "Smith, Laura T. Smith (Laura.Smith2@stjohn.org)" <Laura.Smith2@stjohn.org>;
"lsphysics@att.net" <lsphysics@att.net>
Sent: Friday, August 19, 2016 9:53 AM
Subject: Additional Information request for Bronson Methodist Hospital, NRC License No. 21-13125-01, CN591664

Dear Ms. Smith:

Please see the attached file for additional information needed to complete the review of the recent amendment request concerning the above referenced applicant. Note that the attached letter requests additional information on or before close of business on September 2, 2016. Note that the additional information must be submitted via a dated letter signed a duly authorized management official. Additional guidance may be found in NUREG 1556, Vol. 9, Rev. 2, "Program Program-Specific Guidance About Medical Use Licenses;" which may be found at:

<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v9/r2/>

Submission of your response as a pdf file attached to an email or via facsimile will allow for the quickest processing. Do not hesitate to call me with any questions you may have. Please also send a quick email when you receive this message, to confirm receipt.

Sincerely yours,

Sara A. Forster, Health Physicist Licensing Reviewer
U.S. Nuclear Regulatory Commission - Region III
Division of Nuclear Materials Safety
2443 Warrenville Rd. - Ste. 210
Lisle, IL 60532-4352
sara.forster@nrc.gov
Direct: (630) 829-9892
Facsimile: (630) 515-1078