



Goshen Hospital

September 13, 2016

Materials Licensing Section
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

RE: License #13-18845-01

To whom it may concern:

This is a notification that we will be removing a temporary location of use on our main campus. A mobile PET/CT coach which was operating underneath the canopy outside of the emergency room is to be removed from the license. The last day of operation was September 2, 2016. Only materials identified in 35.200 were used in this area. A closeout survey of the area of use is enclosed.

Please contact our consultant, C. Kelly Stoneberg, at (888) 456-5255 with any questions or concerns.

Sincerely,

Mark Podgorski,
Vice President of Hospital Operations



200 High Park Ave.
PO Box 139
Goshen, IN 46527
574.533.2141

Area Meter Survey

IU GOSHEN PET

Test Date / Time: 09/02/2016 08:43

Instrument: GEIGER COUNTER 2

Make: LUDLUM

Entered By:

Serial#: 219831

ID: 3

Last Calibration Date: 9/2/2016 5:47:48 AM

Background: 0.03 mR/hr

Model: 14C

Instrument check results: 0

Test Data:		Area	Reading	Net results	Action level	Comments
A#	Test location	Type	mR/hr	mR/hr	mR/hr	
1	CLOTHES- TECH 1	Non-Restricted	0.03	0	0.02	
2	HANDS- TECH 1	Non-Restricted	0.03	0	0.02	
3	SHOES-TECH 1	Non-Restricted	0.03	0	0.02	
4	CLOTHES- TECH 2	Non-Restricted	0.03	0	0.02	
5	HANDS- TECH 2	Non-Restricted	0.03	0	0.02	
6	SHOES-TECH 2	Non-Restricted	0.03	0	0.02	
7	CLOTHES- TECH 3	Non-Restricted	0.03	0	0.02	
8	HANDS- TECH 3	Non-Restricted	0.03	0	0.02	
9	SHOES-TECH 3	Non-Restricted	0.03	0	0.02	
10	DOSE CALIBRATOR	Restricted	0.1	0.07	2	
11	HOT SHARPS DISPOSAL	Restricted	0.1	0.07	2	
12	PACKAGE STORAGE CABINET	Restricted	0.3	0.27	2	
13	WASTE DISPOSAL	Restricted	0.03	0	2	
14	PATIENT SCAN TABLE	Restricted	0.03	0	2	on mobile
15	INJECTION CHAIR 1	Restricted	0.03	0	2	
16	INJECTION CHAIR 2	Restricted	0.03	0	2	

Comments And Recommendations

Area Meter Survey

Today is last day for mobile unit. Outgoing survey is:
0.03 mR/Hr—Background
0.03 mR/Hr—Survey

55 cpm Wipe
88 cpm Wipe Background

Performed by: KERR, BRYAN

Radiation safety officer: _____

Forster, Sara

From: Lowden, John <jlowden@goshenhealth.com>
Sent: Wednesday, November 23, 2016 8:32 AM
To: Forster, Sara
Subject: [External_Sender] PET_CT Closeout Letter
Attachments: PET_CT Mobile Removal.pdf; Closeout survey mobile PET 9-16.pdf

If you need anything else let me know. Sorry for the delay I had to rescan the document.

John Lowden
Medical Physicist

CONFIDENTIALITY NOTICE: This transmission and the documents accompanying this transmission may contain confidential information. The information is intended only for the use of the individual(s) or entity named above. If you are not the intended recipient, you are notified that any disclosure, copying, distribution or the taking of any action in reliance on the contents of this information is not permissible. If you have received this message in error, please immediately reply and then destroy it. Thank you.

af