



Deborah Piskura
U.S. NRC
Region III
2443 Warrenville Road
Suite 210
Lisle, IL 60532-4352

Subject:
Report of Source holder stuck open at
8400 Hawthorn Rd.
Kansas City, MO 64120-0013

Deborah:

During a routine shutter check it was discovered that a Berthold source holder with serial number 012/06 was stuck in the open position. The Source holder model # is LB7442-F.

The location of the source was in the SFP Tank Farm at the address listed above in the subject line. It is mounted on the manway of the 1.2B10 Tank. The isotope is a 1000 mCi CS-137 source. The date of the discovery was 10/13/2016. We replaced the existing shield and lockout mechanism with a new shield and lockout mechanism. There was no exposure to any personnel.

Please write or call me at (816) 242-2038 (phone) or (816) 365-2081 (cellphone/voicemail/text) if any further information is needed or email me at steven.scherich@bayer.com.

Sincerely,

A handwritten signature in blue ink, appearing to read "Steve Scherich".

Steven Scherich
Radiation Safety Officer

sss

Bayer
Crop Science Division
Bayer CropScience LP
8400 Hawthorn Road
P.O. Box 4913
Kansas City, MO 64120-0013
816-242-2038
steven.scherich@bayer.com

RECEIVED NOV 22 2016



Bayer CropScience LP
8400 Hawthorne Road
Kansas City, MO 64120

RECEIVED NOV 22 2016

PRESORTED
FIRST CLASS



UNITED STATES POSTAGE
PITNEY BOWES
02 1R
0002011797
\$ 00.45²
NOV 18 2016
MAILED FROM ZIP CODE 66103

U.S.NRC
Attn: Deborah Piskura
Region III
2443 Warrenville Rd.
Suite 210
Lisle, IL 60532-4352

6 JRDONMP 60532

