

**CONNECTICUT HEART GROUP, P. C.**

46 Prince Street, 3rd floor, New Haven, CT 06519  
251 West Main Street, Branford, CT 06405

Br. 1  
03033309

To: USNRC Region I Office  
Licensing Assistance Team  
2100 Renaissance Blvd, Suite 100  
King of Prussia, PA 19406-2713

November 6, 2016

Re: Amendment to License #06-30080-01 for removal of the New Haven, CT site.

Dear Sir or Miss:


This letter serves to request a license amendment for the Connecticut Heart Group, PC. We have stopped the radioactive material delivery and patient imaging at 46 Prince Street, 3rd floor in New Haven, CT. We continue to operate in the office space but we no longer need to use it for nuclear cardiac imaging.

June 20, 2016 was the last day of radioactive material delivery for patient imaging at this office. End of day radiation level surveys and wipe tests for radioactive contamination were conducted and found satisfactory. Our consulting physicist-RSO also performed surveys and an assessment of the suspended service at the site. That data is included on the following page. All of the reference sources (for dose calibrator QA and camera flood source QA) were removed to the remaining Branford office.

If you should have any questions or require additional information regarding our nuclear cardiology imaging office please contact me or Patricia Arnold, Practice Administrator, at 203-315-5300.

We appreciate your efforts to process this amendment and we look forward to a continued safe and effective materials program. Thank you.

Respectfully submitted,



Robert F. Morrison, MD  
President

592373  
RADIOACTIVE MATERIALS-002

**CONNECTICUT HEART GROUP, PC**

46 Prince Street, 3rd floor, New Haven, CT 06519

The following data summarizes the final survey for ambient radiation (mR/hr) levels and removable contamination (dpm) from selected areas of the closed office. There were no areas demonstrating contamination and the final survey results were:

<u>Location</u>	<u>mR/hr</u>	<u>dpm</u>
Imaging Room		
patient table	≤ 0.03	307
gamma camera	≤ 0.03	307
work station	≤ 0.03	307
wipe test equipment	≤ 0.03	307
computer desk	≤ 0.03	307
Hot Lab		
counter	≤ 0.03	274
floor	≤ 0.03	274
Co flood case	1.0	302
Co, Cs sources on contact	2.0	302
Wastes		
general trash	≤ 0.03	274
"hot" boxes	≤ 0.03	274
Floor space, various	≤ 0.03	307
Reception desk & counter	≤ 0.03	307

The background radiation levels were: 0.03 - 0.04 mR/hr (ambient survey), 289 ± 34 dpm (contamination wipe). The survey instruments used were: Ludlum Model 14C GM Survey Meter, and a Ludlum Model 2200 Scaler/Ratemeter and Shielded NaI crystal with a well.



## ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

## Name and Address of Applicant and/or Licensee

Connecticut Heart Group, P.C.  
ATTN: Ajoy Kapoor, M.D., Vice President  
251 West Main Street  
Branford, CT 06405

## Date

November 25, 2016

## License Number(s)

06-30080-01

## Mail Control Number(s)

592373

## Licensing and/or Technical Reviewer or Branch

Medical Branch (Branch 1)

This is to acknowledge receipt of your: ☒ Letter and/or ☐ Application Dated: November 6, 2016

The initial processing, which included an administrative review, has been performed.

☒ Amendment ☐ Termination ☐ New License ☐ Renewal

*Rec'd via fax 11/03/2016*

☒ There were no administrative omissions identified during our initial review.

☐ This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

☐ Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>  
Follow the instructions on the form for submission.

☐ The following administrative omissions have been identified:

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region I  
U. S. Nuclear Regulatory Commission  
Division of Nuclear Materials Safety  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713  
(610) 337-5260, (610) 337-5313,  
(610) 337-5398, or (610) 337-5239