

From: [Lawyer, Dennis](#)
To: scott.coleman@ivsgroup.com
Subject: IVS Hydro, Inc., Request for Additional Information Concerning Application for a New License, Control 592297
Date: Tuesday, December 06, 2016 7:25:00 AM

Dear Scott Coleman,

This is in reference to your application dated October 31, 2016, requesting for a New Nuclear Regulatory Commission License No. 47-35387-01, Docket No. 03039007. In order to continue our review, we need the following additional information:

1. Your application should have been signed by a management representative rather than Scott Coleman, the proposed Radiation Safety Officer. Please submit a letter signed by a management representative indicating that management has reviewed the application and concurs in the statements and representations contained therein. Note also that a management representative should sign all future correspondence that requests a change in your license.
2. Item 5 of your application did not give a specific description of the gauge use. Please provide a specific description of the gauge use.
3. In item 7 of your application you provide the name of the proposed Radiation Safety Officer as Scott Coleman but the only documentation of training and experience was a registration form. Please provide the training certificate of successfully passing the training course.
4. In item 10.2 of your application, you state that you will implement a model radiation survey instrument calibration program. However, you have not requested any sources which would support an instrument calibration program. Please confirm that you plan to support an instrument calibration program or please withdraw this request.
5. In item 10.6 of your application you check off both blocks. Please confirm if the gauge requested does or does not meet any of the safety conditions specified in the "Discussion" part of Section 8.10.6, "Operation, Emergency, and Security Procedures," in NUREG-1556, Volume 4, Revision 1. If it does not, then please provide your operating, emergency, security, and lock-out procedures.
6. In item 10.7 of your application, you checked off that you will send to an outside organization to perform leak tests and that you would implement the model leak test program in Appendix I of NUREG-1556, Volume 4, Revision 1. Please confirm if you will implement a leak test program or send the test out for analysis.
7. We will continue our review upon receipt of this information. Please reply to my attention at the Region 1 Office (Address below) and refer to Mail Control No. 592297. If you have technical questions regarding this letter, please call me at (610) 337-5366.

Your reply must be an originally signed and dated letter. The letter may be scanned and submitted as a pdf document attached to an email; or it may be transmitted by facsimile to

(610) 337-5269; or it may be sent by regular mail. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

Please respond by e-mail to acknowledge that you have received the e-mail request for additional information.

Region 1 Office Mailing Address: Licensing Assistance Team, US Nuclear Regulatory Commission Region I, 2100 Renaissance Boulevard, Suite 100, King of Prussia, PA 19406-2713.

Dennis Lawyer
U.S. NRC Region 1
Health Physicist
610-337-5366