



February 14, 2017

MEMORANDUM TO: File

THROUGH: Aaron T. McCraw, Chief
Materials Inspection Branch 

FROM: Edward Harvey, Health Physicist 

SUBJECT: ATTEMPTED INSPECTION – OAKLAND PHYSICIANS MEDICAL
CENTER, LLC

On January 11, 2017, I attempted to conduct a routine inspection of Oakland Physicians Medical Center in Pontiac, Michigan. The licensee was doing business as Pontiac General Hospital, with authorization to use byproduct materials for diagnostic medical procedures permitted by Title 10 of the *Code of Federal Regulations* Sections 35.100 and 35.200.

At the time of the attempted inspection, the licensee had closed their nuclear imaging department due to flooding caused by a water main break the previous day. As a result, there were no nuclear imaging staff available for inspection activities. After speaking with licensee management, I was granted access to the nuclear imaging department to assess any potential contamination or increased risk to radiological hazards. A tour of the department indicated that a majority of the flood damage was isolated to the camera room. Radioactive material stored in the hot lab appeared to be intact and unaffected by the flood. Independent survey results indicated no signs of contamination.

The licensee's routine inspection cycle remains unchanged as a result of this site visit.

CONTACT: Edward Harvey, MIB
(630) 829-9819

Docket No. 030-02060
License No. 21-06217-02