

CONTROL BLOCK:

--	--	--	--	--	--	--

 ①

(PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

0	1	M	I	D	C	C	1	2	0	0	0	0	0	0	0	0	0	0	0	3	4	1	1	1	1	4			5	
7	8	LICENSEE CODE					14	15	LICENSE NUMBER												25	26	LICENSE TYPE				30	57	CAT	58

CON'T

REPORT SOURCE 01 L 6 0 5 0 0 0 3 1 5 7 1 2 1 8 8 0 8 0 1 1 5 8 1 9

7 8 60 61 DOCKET NUMBER 68 69 EVENT DATE 74 75 REPORT DATE 80

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

02 | DURING A CALIBRATION CHECK OF THE SEISMIC MONITORING PEAK RECORDING

03 | ACCELEROGRAPHS (PRA'S) THE SPENT FUEL POOL, DIESEL GENERATOR ROOM FLOOR

04 | AND CONTAINMENT SPRING LINE PRA'S WERE FOUND TO BE OUT OF CALIBRATION.

05 | THIS CONDITION WAS NONCONSERVATIVE WITH RESPECT TO T.S. 3.3.3.3.

06 | Page

0	7
---	---

08 |

Diagram illustrating the structure of the 20-bit data word:

- SYSTEM CODE: 0 9 (bits 7-8)
- CAUSE CODE: X (bits 9-10)
- CAUSE SUBCODE: Z (bits 11-12)
- COMPONENT CODE: I N S T R U (bits 13-18)
- COMP. SUBCODE: X (bits 19-20)
- VALVE SUBCODE: Z (bits 21-22)

(17) LER/RO REPORT NUMBER	EVENT YEAR [8 0] 21 22		[—] 23	SEQUENTIAL REPORT NO. [0 2 7] 24 25 26		[/] 27	OCCURRENCE CODE [0 3] 28 29		[T] 30	[—] 31	[0] 32
ACTION TAKEN [E] (18)	FUTURE ACTION [X] (19)	EFFECT ON PLANT [Z] (20)	SHUTDOWN METHOD [Z] (21)		HOURS [0 0 0 0] 37 38 39 40	(22)	ATTACHMENT SUBMITTED [N] (23)	NPRD-4 FORM SUB. [Y] (24)	PRIME COMP. SUPPLIER [N] (25)	COMPONENT MANUFACTURER [T 1 0 0] 44 45 46 47	

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

10 | THE CAUSE FOR THE PEAK RECORDING ACCELEROGRAPHS (TELEDYNE GEOTECK,

111 | MODEL PRA-103) BEING OUT OF CALIBRATION IS UNKNOWN AT THIS TIME AND

1 2 | THE INVESTIGATION IS CONTINUING. THE PRA'S WERE RECALIBRATED AND

1 3 | PLACED BACK IN SERVICE. THE ACTION REQUIREMENT OF T.S. 3.3.3.3.a

114 | WAS FULFILLED.

7 8 9
FACILITY STATUS (28) 1 5 9 0 E (28) 10 1 0 0 12 (29) 13 OTHER STATUS (30) NA 44
METHOD OF DISCOVERY (31) B (31) 45 DISCOVERY DESCRIPTION (32) SURVEILLANCE TEST 46 80

ACTIVITY CONTENT
RELEASED OF RELEASE

1 6 Z 33 Z 34

AMOUNT OF ACTIVITY (35)
NA

LOCATION OF RELEASE (36)
NA

PERSONNEL EXPOSURES				DESCRIPTION	
NUMBER			TYPE		
1	7	000	(37) Z	(38)	NA

7		8		9		11		12		13	
PERSONNEL INJURIES											
NUMBER						DESCRIPTION					
1	8	0	0	0	40	NA					

7		8	9	11		12
		LOSS OF OR DAMAGE TO FACILITY				
TYPE		DESCRIPTION		(43)		
1	9	Z	(42)	NA		

ISSUED		PUBLICATION DESCRIPTION	NRC USE ONLY	
2	0	(44) NA	810/200371	
7	8	9	10	80

NAME OF PREPARER STEVE DELONG

PHONE: 616-465-5901

10-11-68