

From: [Lawyer, Dennis](#)
To: harld.w.deason.civ@mail.mil
Subject: Department of the Army, Request for Additional Information Concerning Application for License Renewal, Control 601287, Docket 04009047, Lic. No. STB-1579
Date: Wednesday, November 29, 2017 10:37:00 AM

Dear Mr. Deason,

This is in reference to your application dated September 25, 2017, requesting for renewal to Nuclear Regulatory Commission License No. STB-1579, Docket No. 04009047. In order to continue our review, we need the following additional information:

1. The current license has in the address ATTN: AMSAM-SFO_R. Your application did not list that in your address. Does ATTN: AMSAM-SFO_R need to be placed in the mailing address?
2. Item 10, Instrument Calibration section of your application appears to state that you may use instruments not calibrated with NIST traceable sources for radiation surveys. 10 CFR 20.1501(b) states, "The licensee shall ensure that instruments and equipment used for quantitative radiation measurements (e.g., dose rate and effluent monitoring) are calibrated periodically for the radiation measured." Our guidance states that calibration should be done with NIST traceable sources. Confirm that you will not use these instruments for quantitative radiation measurements.

We will continue our review upon receipt of this information. Please reply to my attention at the Region 1 Office (Address below) and refer to Mail Control No. 601287. If you have technical questions regarding this letter, please call me at (610) 337-5366.

Your reply must be an originally signed and dated letter. The letter may be scanned and submitted as a pdf document attached to an email; or it may be transmitted by facsimile to (610) 337-5269; or it may be sent by regular mail. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application OR amendment request.

Please respond by e-mail to acknowledge that you have received the e-mail request for additional information.

Region 1 Office Mailing Address: Licensing Assistance Team, US Nuclear Regulatory Commission Region I, 2100 Renaissance Boulevard, Suite 100, King of Prussia, PA 19406-2713.

Dennis Lawyer
U.S. NRC Region 1
Health Physicist
610-337-5366