



**UNITED STATES
NUCLEAR REGULATORY COMMISSION**

REGION III
2443 WARRENVILLE RD. SUITE 210
LISLE, IL 60532-4352

October 3, 2017

Anthony Sikorski
Area Manager
InSight Health Corporation
5775 Wayzata Blvd.
STE. 400
Minneapolis, MN 55416

Dear Mr. Sikorski:

In the letter dated August 28, 2017, InSight Health Corporation requested a number of changes to its U.S. Nuclear Regulatory Commission (NRC) license no. 22-29403-01. The request is available electronically from the NRC's Agencywide Documents Access and Management System (ADAMS) at accession no. ML17257A305.

After review of your request, we have noted that additional information is needed to complete our review. Specifically, as discussed with you on October 2, 2017, please provide the following:

A) Request to change the Radiation Safety Officer (RSO):

- 1) Is Ms. Chasteen listed as the RSO on any other NRC or Agreement State licenses? If so, how many licenses is she currently listed on? Please provide the license number, and, if an Agreement State license, a copy of at least one such license with your response;
- 2) How many hours per week will Ms. Chasteen be expected to perform RSO duties for InSight Health Corporation, NRC license no. 22-29403-01?
- 3) Note the amount of time it would be anticipated that the proposed new RSO could respond by phone and in person, in the event of an emergency at one of the temporary job sites;
- 4) Please resubmit the letter, "Delegation of Authority for the Radiation Safety Officer", and include the typed names of individuals signing the letter.

B) In-van use:

- 1) Identify the serial numbers and names of vans that are used within the NRC's jurisdiction;
- 2) Describe the process of radioactive material receipt and use (e.g. inside the van or at the client's facility, is licensee personnel present during receipt of radioactive material etc.);
- 3) Describe the process of radioactive material disposal;
- 4) Describe the secured off-street parking under the licensee's control or the location where the van is parked overnight or when not at the temporary job site;
- 5) Describe licensee's measures to secure the radioactive material within the van to prevent unauthorized access during non-business hours;
- 6) Provide a description of secured storage facilities available for storage of byproduct material and radioactive waste if the van is disabled.

C) Request to remove specific client site locations:

- 1) Describe the radioisotopes used (including unsealed material and sealed sources), quantities, dates of last use, and the name and serial number of the van used at each location;
- 2) Provide history of any spills or leaking sources at each van used at each location;
- 3) Provide a survey map with contamination and exposure rates keyed to specific areas within the van at the location when the material was last used;
- 4) Document the background readings as well as guidance/action levels;
- 5) Describe the instrumentation used including the manufacturer's names and model numbers, last calibration date, and efficiency factors;
- 6) Document the date and name/s of individuals performing the contamination and exposure status survey;
- 7) List the sealed sources used and the leak test results, disposition of each source and receipt acknowledgement letter (from the recipient licensee Radiation Safety Officer) for each source transferred to another authorized recipient, confirming such receipt, if applicable.

D) Confirmation of mailing address on NRC license no. 22-29403-01:


1) Please confirm that the mailing address listed on the license as 5775 Wayzata Blvd., STE. 400, Minneapolis, MN 55416 is current; or

2) If mailing address listed on the license is not current, please provide the mailing address to which the address on the license should be changed.

Please provide a written response to this letter by October 20, 2017. Your response should be signed and dated by an authorized personnel. You may submit your response via facsimile to my attention at 630-515-1078 or as a pdf file attached to an email at Magdalena.Gryglak@nrc.gov. Please reference control no. 600939 in your response.

We will resume our review once we receive your response. If you have any questions, please do not hesitate to contact me at 630-829-9875 or Magdalena.Gryglak@nrc.gov.

Sincerely,



Magdalena R. Gryglak
Health Physicist
Materials Licensing Branch

License No. 22-29403-01

Docket No. 030-38590

Gryglak, Magdalena

From: Gryglak, Magdalena
Sent: Tuesday, October 03, 2017 9:54 AM
To: 'Emily Chasteen'; Tony Sikorski
Subject: Request for Information re NRC License no. 22-29403-01 amendment request dated August 28, 2017
Attachments: RFI.pdf

Mr. Sikorski,

I have attached a request for information letter summarizing the points we discussed yesterday and the information needed to resume our review of your amendment request dated August 28, 2017 for InSight Health Corporation, NRC license no. 22-29403-01. Please contact me if you have any questions regarding the attached document.

Thank you

Magdalena R. Gryglak
U.S. NRC Region III
630-829-9875