



## CONVERSATION RECORD

11/30/17

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU

Anthony Sikorski

DATE OF CONTACT

11/27/2017

TYPE OF CONVERSATION

☐

E-MAIL

☒

TELEPHONE

☐

INCOMING

☐

OUTGOING

E-MAIL ADDRESS

TELEPHONE NUMBER

(952) 513-6806

ORGANIZATION

InSight Health Corporation

DOCKET NUMBER(S)

030-38590

LICENSE NUMBER(S)

22-29403-01

CONTROL NUMBER(S)

CN 600939

SUBJECT

Request for Additional Information

## SUMMARY

On 11/27/17, reviewer contacted Mr. Sikorski for additional information and clarifications regarding the licensee's response dated 11/13/17. See attached email.

**Continue on Page 2**

ACTION REQUIRED (IF ANY)

**Continue on Page 3**

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

## **Gryglak, Magdalena**

---

**From:** Gryglak, Magdalena  
**Sent:** Monday, November 20, 2017 11:17 AM  
**To:** 'Tony Sikorski'  
**Subject:** Amendment Request for InSight Health Corp., NRC License no. 22-29403-01

Mr. Sikorski,

I have a few questions regarding your 11/14/17 response. Please respond to me via email by filling in the missing information and attaching the manufacturer's receipt. I would appreciate if you could respond to me by the end of this week.

**A) Van G331A -City Hospital**

- Please provide the calibration date for the area radiation survey meter, Model 14C, Serial #219367

**B) Van G1309A-Skaggs Outpatient Imaging Center/ Lake Regional Imaging Center**

1) Please provide the calibration date for the area radiation survey meter, Model 14C, Serial #201223 for service on 10/22/14 at the Skaggs Outpatient Imaging Center

2) Please provide the calibration date for the area radiation survey meter, Model 14C, Serial #201223 for service on 11/20/14 at the Lake Regional Imaging Center

3) Please provide the manufacturer's acknowledgement of receipt for Co-57, serial #1193-41-3

4) Please explain the statement for sealed source Ge-68, serial # L6-651 in gantry. Does it mean that you still have the source/PET/CT scanner in your possession?

**C) Van S1351M – Fairmont General Hospital**

1) Please provide the calibration date for the area radiation survey meter, Model 14C, Serial # 190085 for service on 8/23/14 at Fairmont Hospital.

2) Please explain the statement for sealed sources Ge-68, serial # F647, F648, F373 behind gantry? Does it mean that you still have sources/PET/CT in your possession?

**D) Van S1352M (Monongalia General Hospital)**

1) Please provide the calibration date for the area radiation survey meter, Model 14C, Serial # 169962 for service on 11/3/16 at Monongalia General Hospital.

2) Please explain the statement for sealed sources Ge-68, serial # K487, K518, K488 in scanner room? Does it mean that you still have sources in your possession located in the scanner room, in van S1352M?

**E) Van 1334M (Middlesex Hospital/Charlotte Hungerford Hospital)**

1) Please provide the calibration date for the area radiation survey meter, Model 14C, Serial #292558 for service on 7/1/16 and 6/28/16 at both hospitals.

2) Please explain the statement for sealed sources Ge-68, serial # 11005, 19565, 19566 behind gantry, Cs-137, serial #1826-87-10, Co-57, serial #1618-54-7, and 1551-06-22 in hot lab? Does it mean that you still have the sources in your possession located in van 1334M?

3) What does a retired source mean?

Thank you  
Magdalena