

ACCELERATED DISTRIBUTION DEMONSTRATION SYSTEM

REGULATORY INFORMATION DISTRIBUTION SYSTEM (RIDS)

ACCESSION NBR: 9106260130 DOC. DATE: 91/06/21 NOTARIZED: NO DOCKET #
 FACIL: 50-250 Turkey Point Plant, Unit 3, Florida Power and Light C 05000250
 50-251 Turkey Point Plant, Unit 4, Florida Power and Light C 05000251

AUTH. NAME AUTHOR AFFILIATION
 PLUNKETT, T.F. Florida Power & Light Co.
 RECIP. NAME RECIPIENT AFFILIATION
 Document Control Branch (Document Control Desk)

SUBJECT: Provides results of reanalysis of NRC "unknown" sample sent to FPL's Turkey Point Plant on 910308.

DISTRIBUTION CODE: A001D COPIES RECEIVED: LTR 1 ENCL 1 SIZE: 2
 TITLE: OR Submittal: General Distribution

NOTES:

	RECIPIENT ID CODE/NAME	COPIES LTTR ENCL	RECIPIENT ID CODE/NAME	COPIES LTTR ENCL
	PD2-2 LA	1 1	PD2-2 PD	1 1
	AULUCK, R	2 2		
INTERNAL:	NRR/DET/ECMB 7D	1 1	NRR/DET/ESGB	1 1
	NRR/DOEA/OTSB11	1 1	NRR/DST 8E2	1 1
	NRR/DST/SELB 7E	1 1	NRR/DST/SICB8H7	1 1
	NRR/DST/SRXB 8E	1 1	NUDOCS-ABSTRACT	1 1
	OC/LEMB	1 0	OGC/HDS3	1 0
	REG FILE 01	1 1	RES/DSIR/EIB	1 1
EXTERNAL:	NRC PDR	1 1	NSIC	1 1

NOTE TO ALL "RIDS" RECIPIENTS:

PLEASE HELP US TO REDUCE WASTE! CONTACT THE DOCUMENT CONTROL DESK,
 ROOM P1-37 (EXT. 20079) TO ELIMINATE YOUR NAME FROM DISTRIBUTION
 LISTS FOR DOCUMENTS YOU DON'T NEED!

TOTAL NUMBER OF COPIES REQUIRED: LTTR 18 ENCL 16

MA 4



JUN 21 1991

L-91-181

U. S. Nuclear Regulatory Commission
Attn: Document Control Desk
Washington, D. C. 20555

Gentlemen:

Re: Turkey Point Units 3 and 4
Docket Nos. 50-250 and 50-251
Confirmatory Measurements Program

In accordance with Florida Power and Light's (FPL) May 6, 1991, correspondence (L-91-143), and as discussed via telephone between NRC Region II representatives and representatives from Turkey Point's Chemistry and Licensing Departments, this letter provides the results of the reanalysis of the NRC "unknown" sample sent to FPL's Turkey Point Plant March 8, 1991. The reanalysis of the "unknown" provided results that were consistent with the first analysis. However, the Vendor was unable to demonstrate consistency in the results for both sets of National Institute of Standards and Testing (N.I.S.T.) traceable standards.

Because of the consistency in the results for the NRC "unknown" sample, Turkey Point is resubmitting the values for the originally suspect analysis. See attachment 1.

Should there be any questions, please contact us.

Very truly yours,

T. F. Plunkett
Vice President
Turkey Point Nuclear

TFP/DPS/ds

Attachments

cc: Stewart D. Ebnetter, Regional Administrator, Region II, USNRC
Senior Resident Inspector, USNRC, Turkey Point Plant

9106260130 910621
PDR ADOCK 05000250
PDR

2500000

an FPL Group company

A001
11

11

11

11

11

L-91-181
Attachment 1

ATTACHMENT 1

<u>Nuclide</u>	<u>Value ($\mu\text{Ci/ml}$)</u>	<u>% Uncertainties</u>
Fe-55	2.86 E-05	10
Sr-89	4.35 E-05	10

2000
2000

2000

2000

2000

2000