

ACCELERATED DOCUMENT DISTRIBUTION SYSTEM

REGULATOR INFORMATION DISTRIBUTION SYSTEM (RIDS)

ACCESSION NBR: 9407010115 DOC. DATE: 94/06/24 NOTARIZED: NO DOCKET #
 FACIL: 50-250 Turkey Point Plant, Unit 3, Florida Power and Light C 05000250
 50-251 Turkey Point Plant, Unit 4, Florida Power and Light C 05000251
 AUTH. NAME AUTHOR AFFILIATION
 PLUNKETT, T.F. Florida Power & Light Co.
 RECIP. NAME RECIPIENT AFFILIATION
 Document Control Branch (Document Control Desk)

SUBJECT: Forwards results of confirmatory measurements split sample analysis. Results been decay corrected for sample as indicated on attachment.

DISTRIBUTION CODE: A001D COPIES RECEIVED: LTR 1 ENCL 1 SIZE: 2
 TITLE: OR Submittal: General Distribution

NOTES:

	RECIPIENT		COPIES		
	ID CODE/NAME		LTTR	ENCL	
	PD2-2 LA		1	1	
	CROTEAU, R		2	2	
INTERNAL:	NRR/DE/EELB		1	1	NRR/DORS/ONDD
	NRR/DRCH/HICB		1	1	NRR/DRPW
	NRR/DSSA/SPLB		1	1	NRR/DSSA/SPSB
	NRR/DSSA/SRXB		1	1	NUDOCS-ABSTRACT
	OC/LEDCEB		1	0	OGC/HDS3
	<u>REG FILE</u> 01		1	1	
EXTERNAL:	NRC PDR		1	1	NSIC

NOTE TO ALL "RIDS" RECIPIENTS:

PLEASE HELP US TO REDUCE WASTE! CONTACT THE DOCUMENT CONTROL DESK, ROOM P1-37 (EXT. 504-2065) TO ELIMINATE YOUR NAME FROM DISTRIBUTION LISTS FOR DOCUMENTS YOU DON'T NEED!

TOTAL NUMBER OF COPIES REQUIRED: LTTR 17 ENCL 14

MAK



FPL

JUN 24 1994

L-94-164

10 CFR 2.201

U. S. Nuclear Regulatory Commission
Attn: Document Control Desk
Washington, D. C. 20555

Gentlemen:

Re: Turkey Point Units 3 & 4
Docket No. 50-250/251
Confirmatory Measurements Program

This letter forwards the results of the Turkey Point confirmatory measurements split sample analysis. The samples were received on April 25, 1994. The results have been decay corrected for each sample as indicated on the attachment.

Should there be any additional information needed, please contact us.

Very truly yours,

T. F. Plunkett
Vice President
Turkey Point Plant

TFP/CLM/cm

Attachment

cc: Stewart D. Ebnetter, Regional Administrator, Region II, USNRC
Thomas Decker, Section Chief, Radiological Effluents and
Chemistry Section, Region II, USNRC
T. P. Johnson, Senior Resident Inspector, USNRC,
Turkey Point Nuclear Plant

020620
9407010115 940624
PDR ADDCK 05000250
PDR

an FPL Group company

A001
11/1

L-94-164
Attachment

Split Sample Results

The following are the results of the Turkey Point confirmatory measures split samples received on April 25, 1994. The tritium results have been corrected for decay to January 1, 1992; the Sr-90 results, to October 1, 1992; and the Fe-55 results to February 1, 1994. As part of the routine analysis program, QC samples are frequently analyzed with the sample of interest in order to validate the results. This methodology was applied for these split samples.

<u>SAMPLE ID</u>	<u>NUCLIDE</u>	<u>VALUE</u>	<u>% UNCERTAINTIES</u>
NRC 94-2	H-3	3.42 E-05	15
	Fe-55	2.19 E-05	15
	Sr-90	3.86 E-06	15
NRC 94-3	H-3	3.65 E-05	15
	Fe-55	2.44 E-05	15
	Sr-90	3.88 E-06	15
NRC 94-5	H-3	6.18 E-05	15
	Fe-55	2.28 E-05	15
	Sr-90	2.39 E-06	15