

As of: 1/9/18 1:50 PM
 Received: January 08, 2018
 Status: Pending_Post
 Tracking No. 1k2-90t8-i54n
 Comments Due: January 08, 2018
 Submission Type: Web

PUBLIC SUBMISSION

Docket: NRC-2017-0215

Yttrium-90 Microsphere Brachytherapy Sources and Devices Therasphere and SIR-Spheres

Comment On: NRC-2017-0215-0001

Yttrium-90 Microsphere Brachytherapy Sources and Devices TheraSphere and SIR-Spheres;
 Draft Guidance for Comment

Document: NRC-2017-0215-DRAFT-0115

Comment on FR Doc # 2017-24129

82 FR 51655
 11/7/2017

Submitter Information

(114)

Name: David Reindl

Government Agency Type: State

Government Agency: Wisconsin Department of Health Services

General Comment

See attached file(s)

Attachments

STC-17-078 Wisconsin Comments

SUNSI Review Complete
 Template = ADM - 013
 E-RIDS= ADM-03
 Add= Lisa Dimmick (lcdl)



Scott Walker
Governor

Linda Seemeyer
Secretary

State of Wisconsin

Department of Health Services

DIVISION OF PUBLIC HEALTH

1 WEST WILSON STREET
P O BOX 2659
MADISON WI 53701-2659

608-267-4797
FAX: 608-267-3695
TTY: 888-701-1253
dhs.wisconsin.gov

January 8, 2018

Lisa Dimmick
U.S. Nuclear Regulatory Commission
Division of Material Safety, State, Tribal and Rulemaking Programs
Office of Nuclear Material Safety and Safeguards

RE: Comments on FEDERAL REGISTER NOTICE – YTTRIUM-90 (Y-90) MICROSPHERE
BRACHYTHERAPY SOURCES AND DEVICES THERASPHERE® AND SIR-SPHERES®
(STC-17-078)

Dear Ms. Dimmick,

The State of Wisconsin, Radioactive Materials Program has reviewed the above document and submits the following comments:

1. On page 5, in the leak test section, delete the following: "The small size and large number of Y-90 microspheres make leak testing, as required by 10 CFR 35.67(b), impractical. Further, if leak testing were practical, licensees would not be required to leak test individual microspheres."
2. On page 9, the words "in most cases" should be deleted from the last sentence in the first paragraph.
3. On page 9, in the first sentence of the second paragraph, Wisconsin recommends clarifying what "in the process" means. Wisconsin also suggests limiting the continued use of the alternate pathway for physicians who have completed an in-vitro simulated case by the date when this revision to the Y-90 microsphere guidance is published.
4. Wisconsin supports the elimination of the alternate pathway for authorized users.

Sincerely,

David Reindl
Nuclear Engineer
Radioactive Materials Program
State of Wisconsin