



REPLY TO
ATTENTION OF

DEPARTMENT OF THE ARMY
UNITED STATES ARMY AVIATION AND MISSILE COMMAND
5300 MARTIN ROAD
REDSTONE ARSENAL AL 35898-5000

JAN 08 2018

AMSAM-SFO-R

MEMORANDUM FOR Licensing Assistance Team, US Nuclear Regulatory Commission (NRC) Region I, 2100 Renaissance Boulevard, Suite 100, King of Prussia, PA 19406-2713.

ATTENTION: Dennis Lawyer, Health Physicist; Mail Control No. 601287

SUBJECT: Reply to Request for Additional Information Concerning Application for Renewal of NRC License STB-1579 Dated September 25, 2017; Docket No. 04009047

In reply to subject request:

1. The address in the application should include the office code. The complete address would read as follows:

Department of the Army
U.S. Army Aviation and Missile Command
ATTN: AMSAM-SFO-R
5300 Sparkman Center
Redstone Arsenal, Alabama 35898-5000

2. It is confirmed that only instruments calibrated against NIST traceable sources are used for quantitative radiation measurements. Instruments not calibrated to NIST traceable sources are not used for quantitative radiation measurements.

The point of contact for this action is Harold W. Deason, AMCOM Radiation Safety Officer, AMSAM-SFO-R, office 256-842-3250, email harold.w.deason.civ@mail.mil.

Harold W. Deason
Radiation Safety Officer
U.S. Army Aviation and Missile Command
Safety Office

REC'D IN LAT 01/25/2018

601287
PROCESSING MATERIALS-202