

**From:** [Lanzisera, Penny](#)  
**To:** [Frank Digregorio](#)  
**Subject:** Status of submittal of leak tests/disposal records for sealed sources  
**Date:** Thursday, January 11, 2018 3:50:00 PM

---

Molecular Imaging Services, Inc.  
License No. 07-30790-01  
Docket No. 030-36176  
Mail Control 601505

During the last inspection of this site, you possessed 2 dose calibrator sources, 1 flood source and a check source. The record you submitted dated August 21, 2017, documented the transfer of 1 vial (dose calibrator) source and the check source. Please provide the following:

1. Disposal documentation for the 2<sup>nd</sup> dose calibrator source observed during the inspection.
2. Disposal documentation of the flood source observed during the inspection
3. A copy of the PA license
4. Last leak record for any source greater than 100 uCi.

Please submit the above information via fax to 610-337-5269 or via email to my attention and reference Mail Control No. 601505 in your response. Thank you for your assistance,

Penny Lanzisera  
Senior Health Physicist  
U.S. NRC, Region I