

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Rochester Medical Center, P.C.
543 N Main Street
Rochester, MI 48308

REPORT NUMBER(S) 2017001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-32081

4. LICENSE NUMBER(S)

21-26287-01

5. DATE(S) OF INSPECTION

December / 9 2017

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Navid Tehrani		12/12/17
BRANCH CHIEF	Aaron McCraw		02/28/2018

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Rochester Medical Center, P.C. 543 N Main Street Rochester, MI 48308 REPORT NUMBER(S) 2017001		2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352	
3. DOCKET NUMBER(S) 030-32081	4. LICENSE NUMBER(S) 21-26287-01	5. DATE(S) OF INSPECTION December 19, 2017	
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS All		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT Laura Smith, MS - Consultant	4. TELEPHONE NUMBER (734) 455-4730
---------------------------------	----------------------	---	---

☒ Main Office Inspection Next Inspection Date: 12/19/2022

☐ Field Office Inspection _____

☐ Temporary Job Site Inspection _____

PROGRAM SCOPE

This was an unannounced, routine inspection of a private primary care practice authorized to use byproduct material for diagnostic medical purposes at its clinic in Rochester, Michigan. At the time of the inspection, the licensee employed two nuclear medicine technologists who administered general nuclear medicine studies to approximately 500 patients per year. The licensee retained the services of a consulting physicist who audited the radiation safety program on a quarterly basis.

PERFORMANCE OBSERVATIONS

This inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine department in Rochester, and independent measurements. The inspector was unable to observe the administration of any radiopharmaceuticals, as all studies scheduled for the day of the inspection had been completed before the inspector arrived. However, the staff was able to demonstrate the implementation of licensee procedures for receipt of packages containing licensed material and area surveys. The inspector also reviewed a selection of routine nuclear medicine records, as well as in-service training materials and quarterly consultant audits.

No violations of NRC requirements were identified as a result of this inspection.