

## SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:  Capital Cardiology, P.C. 1200 E. Michigan Ave. Lansing, Michigan 48912  REPORT NUMBER(S) 2018001		2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352	
3. DOCKET NUMBER(S)  030-37572	4. LICENSE NUMBER(S)  21-32677-01	5. DATE(S) OF INSPECTION  2/28/18	

## LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

## Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Geoffrey Warren, Sr. HP		2/28/18
BRANCH CHIEF	Aaron McCraw		3/8/18

**Docket File Information****SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

## 1. LICENSEE/LOCATION INSPECTED:

Capital Cardiology, P.C.  
1200 E. Michigan Ave.  
Lansing, Michigan 48912

REPORT NUMBER(S) 2018001

## 2. NRC/REGIONAL OFFICE

Region III  
U. S. Nuclear Regulatory Commission  
2443 Warrenville Road, Suite 210  
Lisle, IL 60532-4352

## 3. DOCKET NUMBER(S)

030-37572

## 4. LICENSE NUMBER(S)

21-32677-01

## 5. DATE(S) OF INSPECTION

February 28, 2018

## 6. INSPECTION PROCEDURES USED

87130

## 7. INSPECTION FOCUS AREAS

03.01 - 03.09

**SUPPLEMENTAL INSPECTION INFORMATION**

## 1. PROGRAM CODE(S)

02201

## 2. PRIORITY

5

## 3. LICENSEE CONTACT

Michelle Kritzman, M.S., RSO

## 4. TELEPHONE NUMBER

(734) 637-2408

☒ Main Office Inspection

Next Inspection Date: 02/28/2023

☐ Field Office Inspection☐ Temporary Job Site Inspection**PROGRAM SCOPE**

This was an unannounced routine inspection. The licensee was a cardiology clinic located in Lansing, Michigan, with authorization to use byproduct materials in 10 CFR 35.200 only. The nuclear medicine department was staffed with one full-time nuclear medicine technologist, who typically administered 120 diagnostic doses monthly, limited to cardiac stress tests. The department received unit doses as needed from a licensed nuclear pharmacy. All waste was either held for decay-in-storage (DIS) or returned to the nuclear pharmacy.

Performance Observations: The inspector observed one diagnostic administration of licensed materials, including dose preparation and disposal. The technologist demonstrated survey meter and wipe counter QC, dose calibrator constancy, package receipt surveys and wipes, daily and weekly contamination surveys, and other procedures. The inspector noted no concerns with these activities. Review of dosimetry records indicated no exposures of regulatory concern. Interviews with licensee personnel indicated adequate knowledge of radiation safety concepts and procedures. The inspector performed independent and confirmatory radiation measurements which indicated results consistent with licensee survey records and postings.

No violations were identified during this inspection.