



**UNITED STATES  
NUCLEAR REGULATORY COMMISSION**  
REGION IV  
1600 E. LAMAR BLVD  
ARLINGTON TX 76011-4511

August 6, 2018

Mr. David P. Tebo  
Corporate Radiation Safety Officer  
TEAM Industrial Services, Inc.  
200 Hermann Drive  
Alvin, TX 77511

SUBJECT: TEAM INDUSTRIAL SERVICES, INC., REQUEST FOR ADDITIONAL  
INFORMATION, MAIL CONTROL NO. 588699

Dear Mr. Tebo:

This is in reference to your application for a license amendment dated September 4, 2015, requesting an exemption to 10 CFR 34.47(a) with the respect to the allowance for radiographers and radiographer's assistants to wear an Instadose dosimeter as their personal dosimeter, regarding NRC License No. 42-32219-01.

1. In order to update condition 24 listed in NRC License No. 42-32219-01, please reply to each of the following items of this correspondence. Your responses must be in the form of a signed and dated letter from someone who has been designated as the licensee certifying officer or management representative for TEAM Industrial Services, Inc.
  - a. How will you identify those Instadose dosimeters needing replacement, recalibration, or a calibration check over the lifetime of use of each Instadose dosimeter you possess?
  - b. What specific criteria will you use for requiring replacement of an Instadose dosimeters in your possession?
  - c. What specific criteria and frequency for calibration checks or recalibrations of dosimeters will you use for Instadose dosimeters in your possession?
2. Additionally, please commit to the following additional requirements.
  - a. That you will only use software provided by Mirion Technologies, Inc., to obtain personnel doses from Instadose dosimeters.
  - b. That in the use of the Instadose dosimeter as a personnel dosimeter you will follow all recommendations by Mirion Technologies, Inc. concerning replacement, recalibration, or a calibration check over the lifetime of use of each Instadose dosimeter you possession.
  - c. That your personnel dosimeter issue periods will be no longer than three months and that measurements are obtained as soon as practical after the issue period ends (i.e., minimize collection to read time).

- d. That all NVLAP-accredited vendor records will be retained per 10 CFR 34.83(c).
- e. That you will maintain records to demonstrate that the criteria in paragraph 1 above had been performed as required.

Please reply to my attention at:

Roberto J. Torres  
Mail Control No. 588699  
U.S. NRC, Region IV  
Division of Nuclear Materials Safety  
1600 East Lamar Boulevard  
Arlington, TX 76011-4511

Please you submit your response to this letter within 30 calendar days from the date of this letter. You can contact me at 817-200-1189 if you have any questions regarding this request for additional information.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

**/RA/**

Roberto J. Torres, M.S., Senior Health Physicist  
Materials Licensing and Inspection Branch

Docket: 030-35252  
License: 42-32219-01  
Control: 588699