

## Gryglak, Magdalena

---

**From:** Gryglak, Magdalena  
**Sent:** Friday, August 10, 2018 6:04 PM  
**To:** 'Evan.Boote@spectrumhealth.org'  
**Subject:** Additional questions, CN 608868, Spectrum Health Hospitals

Good afternoon,

I reviewed the additional information and have additional questions:

- 1) Regarding termination of license no. 21-18667-01, Spectrum Health Pennock, please clarify the discrepancy between the sealed sources listed on the first page of attachment G and the inventory list, in letter dated August 2, 2018.
- 2) Regarding the new layout at the Spectrum Health West Pavilion located at 6105 Wilson Ave SW, Grandville, Michigan, please clarify the purpose of rooms 1027 and 1031, Echo Vascular. Is there any imaging equipment, are these exam rooms, etc.?
- 3) Regarding the release of the former area of use at Spectrum Health West Pavilion and the information provided in your August 9, 2018 letter, please provide the following:
  - specific radioisotopes used
  - the background for the contamination wipes
  - Are the results provided in net dpm?
  - The action levels for ambient radiation surveys and the contamination surveys
  - Confirmation that radioactive waste has decayed to background or was moved to the new area of use
  - Confirmation that all radioactive material was removed from this area of use
  - Confirmation that the sealed sources were moved to the new area of use
  - Sealed source inventory and leak test results performed prior to May 15, 2018
- 4) Please confirm that you do not wish to add James M. Hecksel, D.O. to your license, license no. 21-16737-01.

Please provide information regarding questions 1, 2, and 3 in a signed and dated letter as soon as possible. Please email your response to me (there is no need to mail the letter) so I can quickly process it.

I will be back in the office on Wednesday, August 15. Please let me know if you have any questions.

Thank you

Magdalena R. Gryglak  
U.S. NRC Region III  
630-829-9875

---

**From:** Evan.Boote@spectrumhealth.org [mailto:Evan.Boote@spectrumhealth.org]  
**Sent:** Thursday, August 09, 2018 1:51 PM  
**To:** Gryglak, Magdalena <Magdalena.Gryglak@nrc.gov>  
**Subject:** [External\_Sender] Additional clarification/corrections

Good afternoon Ms. Gryglak,

Attached is a communication in pdf with items that you requested during our phone conversation on August 7. This will also be delivered in hard copy to you via Fedex tomorrow.

Please let me know if you have further questions.

Thank you!

***Evan***

Evan Boote, Ph.D., DABR, FAAPM  
Director, Physics and Technology  
Radiation Safety Officer  
Department of Radiology  
100 Michigan Street  
Grand Rapids, MI 49503  
phone 616.391.2498  
pager 616.479.6175  
mobile 573.424.6905  
[evan.boote@spectrumhealth.org](mailto:evan.boote@spectrumhealth.org)

**SPECTRUM HEALTH**

