

Beaumont

Beaumont Hospital, Royal Oak
3601 West 13 Mile Road
Royal Oak, MI 48073

5 February 2019

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Rd. Suite 210
Lisle, IL 60532-4352

SUBJECT: Amendment to NRC License No. 21-013330-02 Gamma Knife Facility

Dear Ms. Gryglak,


Beaumont Health System is requesting authorization to allow future changes to its radiation safety program for the use of the Icon and will meet the following conditions:

1. The revision is in compliance with the regulations of the NRC or Agreement State;
2. The revision is based on the current guidance for the Icon™ gamma stereotactic radiosurgery unit medical use under 10 CFR 35.1000 use posted on the NRC Web site or the current operators manual and additional safety recommendations from the manufacturer;
3. The revision has been reviewed and approved by the licensee's Radiation Safety Officer and management;
4. The affected individuals are instructed on the revised program before the change is implemented;
5. The licensee will retain a record of each change for 5 years; and
6. The record will include a copy of the appropriate Web site guidance, the old procedure, the new procedure, the effective date of the change, and the signature of the licensee's management representative who reviewed and approved the change.

Should you have any questions, please do not hesitate to contact me.

Thank you in advance for your kind attention in this matter.

Sincerely,



Ann H. Maitz, M.S.

Radiation Safety Officer, Gamma Knife Center
Department of Radiation Oncology
Beaumont Health System – Royal Oak
248-551-1194

Song, Taehoon

From: Gryglak, Magdalena
Sent: Tuesday, February 05, 2019 3:14 PM
To: Sandrik, Lauren; Song, Taehoon; Pavon, Sandy
Cc: Tomczak, Tammy
Subject: FW: Addendum 4
Attachments: Amendment addendum 4 letter.pdf

Please place in ADAMS

Beaumont Health System
030-37359
21-01333-02
CN610456
Thank you

From: Maitz, Ann [mailto:Ann.Maitz@beaumont.org]
Sent: Tuesday, February 05, 2019 2:55 PM
To: Gryglak, Magdalena <Magdalena.Gryglak@nrc.gov>
Subject: [External_Sender] Addendum 4

Hi Magdalena,
Attached please find Addendum 4 specifying the conditions of changes to procedures.