

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

St. Vincent Medical Group, Inc.
8333 Naab Road
Suite 400
Indianapolis, IN 46260

REPORT NUMBER(S) 2019001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-19538

4. LICENSE NUMBER(S)

13-19923-01

5. DATE(S) OF INSPECTION

February 8, 2019

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura, Senior Health Physicist	<i>D. Piskura</i>	2/8/19
BRANCH CHIEF	Aaron T. McCraw, Chief, MIB	<i>Robert B. Gattone, Jr. for ATM</i>	2/20/19

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

St. Vincent Medical Group, Inc.
8333 Naab Road
Suite 400
Indianapolis, IN 46260

REPORT NUMBER(S) 2019001

2. NRC/REGIONAL OFFICE

Region III
U. S. Nuclear Regulatory Commission
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

3. DOCKET NUMBER(S)

030-19538

4. LICENSE NUMBER(S)

13-19923-01

5. DATE(S) OF INSPECTION

February 8, 2019

6. INSPECTION PROCEDURES USED

87130

7. INSPECTION FOCUS AREAS

03.01 - 03.07

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

02201

2. PRIORITY

5

3. LICENSEE CONTACT

Jerry Bishop, RSO

4. TELEPHONE NUMBER

(317) 402-8911

☒ Main Office Inspection

Next Inspection Date: Feb. 8, 2024

☒ Field Office Inspection Greenfield and Lafayette, IN clinics☐ Temporary Job Site Inspection

PROGRAM SCOPE

This was a routine inspection of a large out-patient cardiac clinic authorized to use byproduct material in 10 CFR 35.100 and 35.200 at five locations of use in the central Indiana area. The licensee staffed each of its nuclear medicine departments with one full-time technologist who performed studies daily. Each clinic site administered an average of 50-100 patient studies per month, limited to Tc-99m cardiac procedures. The licensee received its material in unit dose form from a licensed radiopharmacy. The licensee employed a dedicated full-time RSO who performed quarterly audits of the radiation safety program at each clinic site.

This inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine departments at the Greenfield and Lafayette clinics, and independent measurements. All patient studies were completed at the time of the inspector's arrival at the respective locations. Therefore, the inspector could not observe any patient studies, handling of licensed material, etc. The inspection included observations of dose calibrator QA (constancy) checks, security of licensed material, inventories of sealed sources, end of day surveys, package return surveys, and use of personnel monitoring.

No violations of NRC requirements were identified during this inspection.