



Eckert & Ziegler

24937 Avenue Tibbitts
Valencia, California 91355

Isotope Products

Tel 661•309•1010
Fax 661•257•8303

CERTIFICATE OF CALIBRATION GAMMA STANDARD SOURCE

Radionuclide:	Co-60	Customer:	CABRERA SERVICES, INC.
Half-life:	5.272 ± 0.001 years	P.O. No.:	12-0217
Catalog No.:	GF-060-D	Reference Date:	15-May-12 12:00 PST
Source No.:	1592-30-3	Contained Radioactivity:	1.047 µCi 38.74 kBq

Physical Description:

A. Capsule type:	D (25.4 mm OD x 6.35 mm THK)
B. Nature of active deposit:	Evaporated metallic salt
C. Active diameter/volume:	5 mm
D. Backing:	Epoxy
E. Cover:	Acrylic

Radioimpurities:

None detected

Method of Calibration:

This source was assayed using gamma ray spectrometry.

Peak energy used for integration:	1173, 1333 keV
Branching ratio used:	0.9986, 0.9998 gammas per decay

Uncertainty of Measurement:

A. Type A (random) uncertainty:	± 0.5 %
B. Type B (systematic) uncertainty:	± 3.0 %
C. Uncertainty in aliquot weighing:	± 0.0 %
D. Total uncertainty at the 99% confidence level:	± 3.0 %

Notes:

- See reverse side for leak test(s) performed on this source.
- EZIP participates in a NIST measurement assurance program to establish and maintain implicit traceability for a number of nuclides, based on the blind assay (and later NIST certification) of Standard Reference Materials (as in NRC Regulatory Guide 4.15).
- Nuclear data was taken from IAEA-TECDOC-619, 1991.
- This source has a working life of 5 years.

David James Van Delsom
Quality Control

16-Apr-12
Date

EZIP Ref. No.: 1592-30

ISO 9001 CERTIFIED

Medical Imaging Laboratory

24937 Avenue Tibbitts Valencia, California 91355

Industrial Gauging Laboratory

1800 North Keystone Street Burbank, California 91504