



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
3/29/17 10:36:12 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: Daily QC 0329201  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 12:00:00 AM  
Measurement Date: 3/29/17 10:27:14 AM  
Elapsed Live Time: 300.0 seconds  
Elapsed Real Time: 352.1 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.5482E-001+/- 0.011]	9.6116E-001	5.9205E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3252E+000+/- 0.108]	2.2730E+000	-4.8424E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.8910E+000+/- 0.062]	2.0078E+000	1.8723E+000 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6417E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2202E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4085E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.5075E-001+/- 0.008]	5.5162E-001	1.0539E-001 < : : : >
Corrected activity (964. [SD: 4.8373E-001+/- 0.010]	4.8073E-001	-3.0490E-001 < : : : >
Corrected activity (1407 [SD: 4.8383E-001+/- 0.010]	4.9215E-001	8.7068E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)