
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
3/30/17 7:24:50 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: UNC Daily QC 033
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 3/30/17 7:15:33 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 354.5 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.5518E-001+/- 0.010]	9.6991E-001	1.4049E+000 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3223E+000+/- 0.105]	2.4632E+000	1.3393E+000 < : : : >
Peak FWHM (964.13 keV) [SD: 1.8975E+000+/- 0.067]	1.8985E+000	1.4959E-002 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6378E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2195E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4079E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.5080E-001+/- 0.008]	5.6436E-001	1.6964E+000 < : : : >
Corrected activity (964. [SD: 4.8356E-001+/- 0.010]	5.0045E-001	1.7661E+000 < : : : >
Corrected activity (1407 [SD: 4.8429E-001+/- 0.009]	4.9997E-001	1.6550E+000 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)