
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
3/28/17 8:00:01 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0328201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 3/28/17 7:50:28 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 352.3 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.5382E-001+/- 0.010]	9.5189E-001	-1.8766E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3168E+000+/- 0.098]	2.2434E+000	-7.4775E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.8872E+000+/- 0.062]	1.8600E+000	-4.4035E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6387E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2197E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4080E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4910E-001+/- 0.006]	5.5326E-001	6.6753E-001 < : : : >
Corrected activity (964. [SD: 4.8228E-001+/- 0.009]	4.8543E-001	3.5327E-001 < : : : >
Corrected activity (1407 [SD: 4.8229E-001+/- 0.008]	4.8529E-001	3.6226E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)