



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
6/21/17 8:30:15 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: Daily QC 0621201  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 2:00:00 PM  
Measurement Date: 6/21/17 8:19:47 AM  
Elapsed Live Time: 300.0 seconds  
Elapsed Real Time: 351.4 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6393E-001+/- 0.014]	9.5622E-001	-5.7109E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3894E+000+/- 0.144]	2.3592E+000	-2.0921E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9197E+000+/- 0.086]	1.8990E+000	-2.3994E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6447E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2201E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4089E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4630E-001+/- 0.012]	5.4937E-001	2.6471E-001 < : : : >
Corrected activity (964. [SD: 4.8364E-001+/- 0.007]	4.8210E-001	-2.1489E-001 < : : : >
Corrected activity (1407 [SD: 4.8505E-001+/- 0.007]	4.8133E-001	-5.4720E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)