
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
6/26/17 12:34:16 PM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 062620
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 2:00:00 PM
Measurement Date: 6/26/17 12:27:02 PM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.3 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6411E-001+/- 0.013]	9.6421E-001	7.7067E-003 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3888E+000+/- 0.141]	2.2654E+000	-8.7582E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9201E+000+/- 0.084]	1.7965E+000	-1.4781E+000 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6388E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2191E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4080E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4653E-001+/- 0.011]	5.4444E-001	-1.8519E-001 < : : : >
Corrected activity (964. [SD: 4.8372E-001+/- 0.007]	4.8679E-001	4.3609E-001 < : : : >
Corrected activity (1407 [SD: 4.8509E-001+/- 0.007]	4.8146E-001	-5.4373E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)