



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
8/28/17 9:40:33 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: Daily QC 0825201  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 2:00:00 PM  
Measurement Date: 8/28/17 9:14:02 AM  
Elapsed Live Time: 600.0 seconds  
Elapsed Real Time: 701.9 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6411E-001+/- 0.013]	9.6048E-001	-2.7523E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3870E+000+/- 0.141]	2.2350E+000	-1.0816E+000 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9183E+000+/- 0.084]	1.8344E+000	-9.9455E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6429E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2197E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4087E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4650E-001+/- 0.011]	5.4290E-001	-3.2058E-001 < : : : >
Corrected activity (964. [SD: 4.8376E-001+/- 0.007]	4.8740E-001	5.2087E-001 < : : : >
Corrected activity (1407 [SD: 4.8504E-001+/- 0.007]	4.8355E-001	-2.2400E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)