
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
6/8/17 9:32:58 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0608201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 2:00:00 PM
Measurement Date: 6/8/17 8:21:50 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.5 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6240E-001+/- 0.011]	9.7782E-001	1.3613E+000 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3898E+000+/- 0.145]	2.4685E+000	5.4256E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9188E+000+/- 0.087]	1.9670E+000	5.5180E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6404E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2193E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4083E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4684E-001+/- 0.012]	5.4568E-001	-1.0114E-001 < : : : >
Corrected activity (964. [SD: 4.8384E-001+/- 0.007]	4.8403E-001	2.8111E-002 < : : : >
Corrected activity (1407 [SD: 4.8528E-001+/- 0.007]	4.8013E-001	-7.6112E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)