
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
6/15/17 8:20:13 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0615201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 2:00:00 PM
Measurement Date: 6/15/17 8:00:15 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.2 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6266E-001+/- 0.011]	9.7837E-001	1.3774E+000 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3911E+000+/- 0.144]	2.2538E+000	-9.5150E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9196E+000+/- 0.087]	1.8544E+000	-7.5240E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6419E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2196E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4085E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4682E-001+/- 0.011]	5.4358E-001	-2.8327E-001 < : : : >
Corrected activity (964. [SD: 4.8384E-001+/- 0.007]	4.7638E-001	-1.1108E+000 < : : : >
Corrected activity (1407 [SD: 4.8519E-001+/- 0.007]	4.8497E-001	-3.2640E-002 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)