



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
6/16/17 7:43:38 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: Daily QC 0616201  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 2:00:00 PM  
Measurement Date: 6/16/17 7:31:09 AM  
Elapsed Live Time: 300.0 seconds  
Elapsed Real Time: 351.8 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6292E-001+/- 0.011]	9.6752E-001	4.0088E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3888E+000+/- 0.144]	2.6419E+000	1.7558E+000 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9186E+000+/- 0.086]	2.0645E+000	1.6892E+000 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6377E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2190E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4078E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4677E-001+/- 0.011]	5.4664E-001	-1.1181E-002 < : : : >
Corrected activity (964. [SD: 4.8372E-001+/- 0.007]	4.8522E-001	2.2243E-001 < : : : >
Corrected activity (1407 [SD: 4.8519E-001+/- 0.007]	4.9302E-001	1.1719E+000 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)