
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
6/20/17 10:46:29 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 062020
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 2:00:00 PM
Measurement Date: 6/20/17 10:28:28 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 349.4 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6388E-001+/- 0.013]	9.4743E-001	-1.2345E+000 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3910E+000+/- 0.146]	2.3179E+000	-5.0014E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9212E+000+/- 0.087]	1.8550E+000	-7.6002E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6409E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2193E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4084E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4630E-001+/- 0.012]	5.4285E-001	-2.9360E-001 < : : : >
Corrected activity (964. [SD: 4.8341E-001+/- 0.007]	4.8846E-001	7.0621E-001 < : : : >
Corrected activity (1407 [SD: 4.8507E-001+/- 0.007]	4.8488E-001	-2.7623E-002 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)