
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
5/15/17 8:23:11 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 051520
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 5/15/17 8:10:42 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.7 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6189E-001+/- 0.012]	9.5686E-001	-4.3277E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3914E+000+/- 0.142]	2.3082E+000	-5.8665E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9185E+000+/- 0.084]	1.8899E+000	-3.4268E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6420E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2196E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4085E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4713E-001+/- 0.012]	5.4590E-001	-1.0017E-001 < : : : >
Corrected activity (964. [SD: 4.8424E-001+/- 0.007]	4.8811E-001	5.5682E-001 < : : : >
Corrected activity (1407 [SD: 4.8504E-001+/- 0.007]	4.9008E-001	7.0922E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)