
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
5/16/17 8:43:18 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0516201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 5/16/17 8:32:08 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.7 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6180E-001+/- 0.012]	9.6206E-001	2.2967E-002 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3898E+000+/- 0.141]	2.4693E+000	5.6382E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9180E+000+/- 0.083]	2.0284E+000	1.3325E+000 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6421E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2197E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4086E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4710E-001+/- 0.012]	5.4275E-001	-3.5903E-001 < : : : >
Corrected activity (964. [SD: 4.8431E-001+/- 0.007]	4.8194E-001	-3.4437E-001 < : : : >
Corrected activity (1407 [SD: 4.8513E-001+/- 0.007]	4.8838E-001	4.5885E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)