
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
5/23/17 9:25:59 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0523201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 5/23/17 9:18:31 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.4 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6211E-001+/- 0.012]	9.5672E-001	-4.6668E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3878E+000+/- 0.141]	2.2756E+000	-7.9538E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9174E+000+/- 0.085]	1.8528E+000	-7.6000E-001 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6417E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2195E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4085E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4700E-001+/- 0.012]	5.4371E-001	-2.7563E-001 < : : : >
Corrected activity (964. [SD: 4.8415E-001+/- 0.007]	4.8165E-001	-3.6519E-001 < : : : >
Corrected activity (1407 [SD: 4.8525E-001+/- 0.007]	4.8756E-001	3.3141E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)