
***** G E N I E Q U A L I T Y A S S U R A N C E *****

Last Results Report
5/3/17 8:00:53 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856
Sample ID: Daily QC 0503201
Sample Quantity: 1.0000E+000
Sample Date: 6/1/11 12:00:00 AM
Measurement Date: 5/3/17 7:52:58 AM
Elapsed Live Time: 300.0 seconds
Elapsed Real Time: 351.7 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6196E-001+/- 0.012]	9.5854E-001	-2.8221E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3869E+000+/- 0.139]	2.4426E+000	4.0096E-001 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9187E+000+/- 0.085]	1.9256E+000	8.1808E-002 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6426E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2197E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4086E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4918E-001+/- 0.006]	5.4169E-001	-1.1973E+000 < : : : >
Corrected activity (964. [SD: 4.8456E-001+/- 0.007]	4.8401E-001	-7.6975E-002 < : : : >
Corrected activity (1407 [SD: 4.8495E-001+/- 0.007]	4.8031E-001	-6.7672E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)
BS = Measurement Bias Test (In = Investigate, Ac = Action)