



\*\*\*\*\*  
\*\*\*\*\* G E N I E Q U A L I T Y A S S U R A N C E \*\*\*\*\*  
\*\*\*\*\*

Last Results Report  
5/4/17 7:18:25 AM

QA File: C:\GENIE2K\CAMFILES\UNC 2017 856  
Sample ID: Daily QC 0504201  
Sample Quantity: 1.0000E+000  
Sample Date: 6/1/11 12:00:00 AM  
Measurement Date: 5/4/17 7:10:51 AM  
Elapsed Live Time: 300.0 seconds  
Elapsed Real Time: 351.8 seconds

Parameter Description [Mean +/- Std. Dev.]	Value	Deviation/Flags < LU : SD : UD : BS >
Peak FWHM (121.8 keV) [SD: 9.6189E-001+/- 0.012]	9.5916E-001	-2.2689E-001 < : : : >
Peak FWHM (1407.92 keV) [SD: 2.3882E+000+/- 0.137]	2.6177E+000	1.6698E+000 < : : : >
Peak FWHM (964.13 keV) [SD: 1.9188E+000+/- 0.084]	2.0178E+000	1.1760E+000 < : : : >
Peak energy (964.13 keV) [0.00+/-0.00]	9.6411E+002	0.00 < : : : >
Peak energy (121.8 keV) [0.00+/-0.00]	1.2195E+002	0.00 < : : : >
Peak energy (1407.92 keV) [0.00+/-0.00]	1.4084E+003	0.00 < : : : >
Corrected activity (121. [SD: 5.4902E-001+/- 0.006]	5.4270E-001	-1.0054E+000 < : : : >
Corrected activity (964. [SD: 4.8455E-001+/- 0.007]	4.8612E-001	2.2128E-001 < : : : >
Corrected activity (1407 [SD: 4.8485E-001+/- 0.007]	4.8218E-001	-3.9170E-001 < : : : >

Flags Key: LU = Lower/Upper Bounds Test (Ab = Above, Be = Below)  
SD = Sample Driven N-Sigma Test (In = Investigate, Ac = Action)  
UD = User Driven N-Sigma Test (In = Investigate, Ac = Action)  
BS = Measurement Bias Test (In = Investigate, Ac = Action)